



Doc No: CC/EL120/14.03.2022

AGD Biomedicals Pvt. Ltd
AGD Biomedicals (P) Ltd. Mehta
Trade Center, No.1, Sir Mathuradas
Vasanji Rd, Shivaji Colony, Andheri
East, Mumbai, Maharashtra 400059

Calibration Certificate

Customer Name : Antlia Lab and Diagnostics Pvt. LTD Ranchi

Model Name : EL-120 Electrolyte Analyzer

Serial Number :BD211002

Calibration Date :14.03.2023

Validity of Calibration : 13.03.2024

Authorized Signatory Rakesh Kumar

AGD Biomedicals (P) Ltd

Note :Recommended to perform calibration every 6 months.

Calibration Passed

Model Name : EL 120
Date of Performance :14.03.2022

Calibration Performance										
Parameters	S1	S2	S3	Mean	SD	CV	Target	Range	1 SD	2 SD
Sodium	141	141	141	141.00	0.00	0.00	142.00	140-144	3.00	6.00
Potassium	4.7	4.69	4.65	4.68	0.03	0.57	4.68	4.65-4.71	0.20	0.40
Chloride	101	99	100	100.00	1.00	1.00	100.00	98-102	0.15	0.30

Name and Signature of Customer

Seal & Date

Name and Signature of Engineer

Seal & Date