



Doc No: CC/EL120/14.03.2022

AGD Biomedicals Pvt. Ltd
AGD Biomedicals (P) Ltd. Mehta
Trade Center, No.1, Sir Mathuradas
Vasanji Rd, Shivaji Colony, Andheri
East, Mumbai, Maharashtra 400059

Calibration Certificate

Customer Name : Antlia Lab and Diagnostics Pvt. LTD Ranchi

Equipment Name : Clinical Chemistry Analyser

Model Name : AGD2020 Clinical Chemistry Analyser

Serial Number : AB2109039

Calibration Date : 14.03.2023

Validity of Calibration : 13.03.2024

Authorized Signatory Rakesh Kumar

AGD Biomedicals (P) Ltd

Note : Recommended to perform calibration every 6 months.

Calibration Passed

Model Name : AGD 2020 Clinical Chemistry Analyser
Date of Performance : 14.03.2023

Calibration Performance										
Parameters	S1	S2	S3	Mean	SD	CV	Target	Range	1 SD	2 SD
Glucose	99	99	100	99.33	0.58	0.58	100.00	98-102	1.00	1.00
Tryglycerides	149.9	149	150	149.63	0.55	0.37	150.00	148-152	1.00	2.00
SGOT	50	50	51	50.33	0.58	1.15	50.00	48-52	1.00	2.00
Creatinine	2.2	1.99	2.2	2.13	0.12	5.69	2.20	2.15-2.25	0.25	0.25

Shubham Choudhary

Name and Signature of Customer
Seal & Date



Rakesh

Name and Signature of Engineer
Seal & Date

