



**PROFICIENCY TESTING REPORT,  
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
NABL accredited program as per ISO/IEC 17043:2010 standard  
Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 4317

Distribution No.: 156-K

Month/Year: July/2022

Instrument ID: 705ES0H11486

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 25-09-2022[Final].

**CBC and Retic Assessment**

Test Parameters	S.No.	Among Lab (Accuracy Testing)					Within Lab (Precision Testing)				
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC $\times 10^3/\mu\text{l}$	1	5.6	5.5	11.1	9.2	0.0610	1.22	0.1	0.1	0.0110	0.00
RBC $\times 10^6/\mu\text{l}$	1	4.38	4.29	8.67	8.88	0.0110	-0.79	0.09	0.05	0.0030	0.77
Hb g/dl	1	13.6	13.2	26.8	27	0.0370	-0.21	0.4	0.1	0.0100	2.02
HCT%	1	45.4	44.1	89.5	83.2	0.2210	1.17	1.3	0.5	0.0370	1.54
MCV-fL	1	104	103	207	186.65	0.3960	2.26	1	0.3	0.0270	1.57
MCH-Pg	1	31	30.7	61.7	60.9	0.0870	0.41	0.3	0.3	0.0190	0.00
MCHC-g/dl	1	29.9	29.9	59.8	65.15	0.1890	-1.22	0	0.3	0.0210	-0.95
Plt. $\times 10^3/\mu\text{l}$	1	174	149	323	211.5	1.74	2.40	25	4	0.32	4.72
Retic %	2	3.7	3.5	7.2	3.8	0.10	1.36	0.2	0.2	0.01	0.00

**P.S . Assesment**

YOUR REPORT			CONSENSUS REPORT
DLC%	3	Nrbc=1 , Poly=40 L=2, E=2, Mono/Promono=1 , B1=1 P.M.=2, Mve=42. Meta=7. Other=	Poly: 35 - 48, Myelo: 14 - 26, Meta: 8 - 15, Promyelo: 3-7, nRBC/ Lympho /Blast/Eos/Baso/Mono: 0 - 5
RBC Morphology	3	RBC SHOW MILD ANISOCYTOSIS,PREDOMINANTLY NORMOCYTIC NORMOCHROMIC, MICROCYTES(+),MACROCYTES(+) SEEN.	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis
Diagnosis	3	MYELOPROLIFERATIVE NEOPLASM, MOST PROBABLY CHRONIC MYELOID LEUKEMIA-CHRONIC PHASE (CML-CP)	Chronic Myeloid Leukemia (Chronic Phase)

### COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 156-K	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC $\times 10^3/\mu\text{l}$	1	252	248	84.27	87.1	6.85	2.42	8.88	10.48
RBC $\times 10^6/\mu\text{l}$	1	252	252	86.11	92.46	6.75	1.59	7.14	5.95
Hb g/dl	1	252	252	90.87	82.54	1.98	7.94	7.15	9.52
HCT%	1	252	248	94.35	91.94	2.42	2.82	3.23	5.24
MCV-fL	1	252	248	89.11	94.35	6.85	2.42	4.04	3.23
MCH-Pg	1	252	248	87.1	91.53	6.05	2.02	6.85	5.65
MCHC-g/dl	1	252	248	91.13	91.94	4.44	2.42	4.43	5.64
PLT $\times 10^3/\mu\text{l}$	1	252	248	92.34	89.11	3.63	5.24	4.03	5.65
ReticCount%	2	252	214	93.46	89.72	4.67	7.48	1.87	2.80
PS Assessment	3	252	213	Satisfactory : 93.65%, Borderline Sat. : 4.76%, Unsatisfactory : 1.58%					

**Comments:**

- 1). Among Lab (EQA) : Results acceptable.
- 2). Within Lab (IQA) : Difference in the CBC measurement values for PLT unacceptable, may be due to random/human error.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR =  $0.7413 \times \text{IQR}$

**Note-3:** Z score 0 to +2 : Acceptable, Z score +2 to +3 : Warning Signal, Z score > +3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3 \times \text{SDPA}$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x} - \bar{y}$ ) should be smaller than the check value ( $0.3 \times \text{SDPA}$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,

Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029

*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

**EQAP CODE No. : 4317**

**Distribution No.: 158-K**

**Month/Year: December/2022**

**Instrument ID: MICROS ES60 (705ESOH11486)**

**Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi.**  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

**Date of issue & status of the report: 23-02-2023[Final].**

**CBC and Retic Assessment**

<b>Test Parameters</b>	<b>S.No.</b>	<b>Among Lab (Accuracy Testing)</b>					<b>Within Lab (Precision Testing)</b>				
		<b>Your Result 1</b>	<b>Your Result 2</b>	<b>Your Results Sum of 2 Value</b>	<b>Consensus result sum of 2 values (Assigned Value)</b>	<b>Uncertainty of Assigned Values</b>	<b>Z Score</b>	<b>Yours Results Diff. of 2 Values</b>	<b>Consensus Result Diff. of 2 values (Assigned Value)</b>	<b>Uncertainty of Assigned Values</b>	<b>Z Score</b>
<b>WBC <math>\times 10^3/\mu\text{l}</math></b>	1	4.7	4.4	9.1	9.5	0.0430	<b>-0.40</b>	0.3	0.1	0.0290	<b>1.93</b>
<b>RBC <math>\times 10^6/\mu\text{l}</math></b>	1	3.61	3.59	7.2	7.47	0.0090	<b>-1.40</b>	0.02	0.03	0.0030	<b>-0.34</b>
<b>Hb g/dl</b>	1	11.7	11.6	23.3	24	0.0250	<b>-1.18</b>	0.1	0.1	0.0080	<b>0.00</b>
<b>HCT%</b>	1	35.4	34.9	70.3	73	0.1780	<b>-0.60</b>	0.5	0.3	0.0270	<b>0.67</b>
<b>MCV-fL</b>	1	99	97	196	196.1	0.3950	<b>-0.01</b>	2	0.3	0.0290	<b>3.82</b>
<b>MCH-Pg</b>	1	32.4	32.3	64.7	64.2	0.0830	<b>0.26</b>	0.1	0.3	0.0180	<b>-0.90</b>
<b>MCHC-g/dL</b>	1	33.5	32.7	66.2	65.5	0.1560	<b>0.17</b>	0.8	0.3	0.0220	<b>1.69</b>
<b>Plt. <math>\times 10^3/\mu\text{l}</math></b>	1	198	196	394	312	1.36	<b>2.77</b>	2	4	0.34	<b>-0.45</b>
<b>Retic %</b>	2	6.2	6.1	12.3	8	0.19	<b>0.89</b>	0.1	0.3	0.02	<b>-0.90</b>

**P.S . Assesment**

<b>YOUR REPORT</b>			<b>CONSENSUS REPORT</b>
<b>DLC%</b>	3	Nrbcs=3 , Poly=60 L=12, E=6, Mono/Promono=4 , B1=0 P.M.=1, Mye=14, Meta=1, Other=	Poly: 51 - 65, Myelo: 5 - 12, Meta: 5- 11, Lympho: 7- 14, Eosino: 2-6, Promyelo: 0-5, nRBC/Blast/Baso/Mono: 0 - 5
<b>RBC Morphology</b>	3	PREDOMINANTLY NORMOCYTIC NORMOCHROMIC WITH MICROCYTE (+) AND MACROCYTE (+)	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis
<b>Diagnosis</b>	3	1. MYELOPROLIFERATIVE DISORDER SUGGESTIVE OF CHRONIC MYELOID LEUKEMIA - CHRONIC PHASE (CML-CP) 2. NEUTROPHILIC LEUCOCYTOSIS WITH LEFT SHIFT	Chronic Myeloid Leukemia (Chronic Phase)

### COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 158--K	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC $\times 10^3/\mu\text{l}$	1	233	229	83.41	88.65	1.75	5.24	14.84	6.11
RBC $\times 10^6/\mu\text{l}$	1	233	233	83.26	85.84	8.15	7.3	8.59	6.86
Hb g/dl	1	233	233	88.41	91.42	6.87	0.86	4.72	7.72
HCT%	1	233	233	94.32	88.65	4.37	3.93	1.31	7.42
MCV-fL	1	233	229	97.38	97.38	2.18	0.87	0.44	1.75
MCH-Pg	1	233	229	90.83	91.27	6.11	4.37	3.06	4.36
MCHC-g/dl	1	233	229	97.38	87.77	1.31	6.55	1.31	5.68
Plt. $\times 10^3/\mu\text{l}$	1	233	229	90.39	93.45	7.86	5.24	1.75	1.31
ReticCount%	2	233	204	96.08	86.27	3.43	1.96	0.49	11.77
PS Assessment	3	233	198	Satisfactory :87.56%, Borderline Sat. :10.30, Unsatisfactory :2.14%					

**Comments:**

- 1). Among Lab (EQA) : Results acceptable.
- 2). Within Lab (IQA) : Difference in the CBC measurement values for MCV unacceptable, may be due to random/human error.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR =  $0.7413 \times \text{IQR}$

**Note-3:** Z score 0 to  $\pm 2$  : Acceptable, Z score  $\pm 2$  to  $\pm 3$  : Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3 \times \text{SDPA}$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x} - \bar{y}$ ) should be smaller than the check value ( $0.3 \times \text{SDPA}$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishhtmaiimseqap.com](http://www.ishhtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,

Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----