



Customer Name & Address : M/s GOVERNMENT PRIMARY HEALTH CENTRE GAJALNAYAKANPATTI, THIRUPATTUR DT	Report No	VCPL-22-23-CN-115/06
	Date of Receipt	21-Mar-2023
	Date of Issue	22-Mar-2023
	Calibration Date	21-Mar-2023
	Next Due Date	20-Mar-2024
Customer Reference : Telephone Communication	Calibration at	On Site

Details of Device Under Calibration:

Description:	SEMI AUTO ANALYZER	Make:	ROBONIK
Range:	-	Model/Type:	PRIETEST TOUCH
Resolution:	-	Condition of DUC:	Good
Serial Number:	ATCD2970818RBK	Location:	LABORATORY
ID number:	-	Cal. At:	ONSITE

Environmental Conditions		Calibration Method Used	
Temperature	25.9°C	National/International standard:	IEC60601-1
Humidity	52 %Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyser	05H-0600	ACCS221497	Rigel Medical / 288+	9-Aug-23

Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
1	Earth Leakage	152	<5000 μ A For B,BF,CF
		MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
2	Enclosure Leakage	210	<500 μ A For B,BF,CF

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


C.AKASH

CALIBRATION ENGINEER



Authorised by:


S. LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

*** End of the Report ***

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indira Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

Web : vanavilcalibrations.com | Email : vanavilcalibrations@gmail.com