

### CALIBRATION CERTIFICATE

CUSTOMER: LIFETECH DIAGNOSTIC, SALEM.

Unit Description: Microscope

Make: LABOMED

Room Temperature :25°C ± 2 Room Humidity :30-50 % (At the time of calibration)

#### ESA Test Result


S.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1.	Current	90	Ma	1	Normal
2.	Frequency	50.0	Volts	49-50	Normal
3.	Mains voltage line-Neutral	230.3	Volts	220-240	Normal
4.	Mains voltage line-Earth	230.3	Volts	220-240	Normal
5.	Mains voltage Neutral-Earth	4.8	Volts	<7	Normal


S.No	Knob Position	DUC Reading (LUX)	Distance (Cm)
1.	Low	3270	30
2.	High	19110	30

CALIBRATION DATE: 01.03.2023

CALIBRATION VALIDITY: 31.03.2024

TEST BY: G. DINESH

APPROVED BY: 



NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.