

Ph. : (O) + 91 265 3927448/49/50
2632148/49/50
Fax : (O) + 91 265 2632151

LABLINE EQUIPMENTS PVT. LTD.

An ISO 9001 : 2008 & 13485: 2003 Accredited Company



Mfg. Of Medical Equipments

614/1, G.I.D.C. B/H Himalaya Machinery, Makarpura, Vadodara - 390 010, Gujarat, INDIA
E-mail : haresh@lablineindia.com • mail@lablineindia.com • Web Site: www.lablineindia.com

CALIBRATION CERTIFICATE

CUSTOMER :

LIFETECH DIAGNOSTICS
234/3, Sri Nagar Colony,
Narasothipatti, Kuranguchavadi,
Salem - 636 004

CERTIFICATE NO. **LL/257/20-21**

DEALER :

Labtech Medico (P) Ltd
XI-434 Chakkiath Joseph Memorial Building,
Parappuram Junction, Karukutty P.O,
Ernakulam District, Kerala-683576.

DESCRIPTION OF ITEM : Digital Medical Centrifuge (CF 655) with Rotor RT 07 (16 x 15 ml Swing out Head)
CALIBRATION DATE : 08/08/2022

MODEL NO.: CF 655 M/C.Sr.No. 594 - A

RANGE : 1000 - 3500 RPM

DETAILS OF EQUIPMENTS USED FOR CALIBRATION :--

SR. NO.	DESCRIPTION	CERTIFIED BY	CERTIFICATE NO.	VALIDITY
1	DIGITAL TACHOMETER	G.I.I.	I - 3776	1 YEAR

OBSERVATION :

R.P.M. CALIBRATION * [For Centrifuges only]

POINT	INSTRUMENT R.P.M. DISPLAYED	CALIBRATED TACHO METER READING	ERROR OBSERVED	ACCEPTED ERROR	RESULT
1	1000	1051	-51	+/- 100	OK
2	2500	2537	-37	+/- 100	OK
3	3500	3542	-42	+/- 100	OK

TEMPERATURE CALIBRATION * : [For Dry bath incubators only]

POINT	INSTRUMENT TEMPERATURE DISPLAYED (°C)	CALIBRATED THERMOMETER TEMP. (°C)	ERROR OBSERVED	ACCEPTED ERROR	RESULT

NEXT CALIBRATION DUE ** :

APPROVED BY:
Labline Equipments Pvt. Ltd.

Director
Labline Equipments Pvt. Ltd.
614/1, G.I.D.C. Estate,
Makarpura, Vadodara - 390 010.
Phone: 3927448, 49, 50.

INSPECTION DEPARTMENT

* : RPM & temperature readings are taken on "LABLINE EQUIPMENTS PVT. LTD." standard testing panels and instruments. All readings are taken under ideal test condition.

** : As required by the user.