



UNMATCHED SERVICE
SINCE 1979...

Date:
Effective Date:

Certificate of Calibration

Customer Name: Regional Ayurveda Research Institute

Model : Automated Hematology Analyzer Sysmex XP-100

Serial No. : A1143

Calibration Done Date: 17.06.2022

Next Calibration Due Date On or Before: 16.06.2023

Lab In-charge: . Sangeeta Ramteke

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Sysmex Corporation, Japan.

The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.



Calibration at site performed by
Engineer Name
Designation
Transasia Bio-Medicals Ltd
Location

Encl:

1. Certificate of Inspection
2. Assay Sheet of Calibrator SCS-1000
3. Printouts
4. Traceability & Uncertainty document



Date: 00-01-1900
Effective Date: 00-01-1900

UNMATCHED SINCE 1978

Certificate of Inspection

1. Model: Automated Hematology Analyzer Sysmex XP – 100
2. Serial No.: A1143
3. Calibration Date: 17.06.2022
4. Material used: SCS-1000 (Lot No. 21370525, Expiry date: 19-June-2022)

By comparing your data to the results of the standard counters in Sysmex Corporation, the calibration for CBC 5 parameters using the measurement standard material (SCS-1000) was completed. The calibration result of 5 runs is summarized in the following table. Please refer to the attached sheets for the details.



Technical Service Department
Transasia Bio-Medicals Ltd