



Date: 11/02/2023
Effective Date: 11/02/2023

Certificate of Calibration

Customer Name: Global Pathology Center, Hari Nagar, New Delhi

Model : Automated Hematology Analyzer Sysmex XP100

Serial No. : A-6927

Calibration Done Date: 11/02/2023

Next Calibration Due Date On or Before: 12/02/2024

Lab In-charge: . Dr.R.S.Khushwaha

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Sysmex Corporation, Japan.

The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.

Calibration at site performed by
Engineer Name: Kumar Saurabh
Designation : Service Engineer
Transasia Bio-Medicals Ltd
Location:New Delhi

Encl:

1. Certificate of Inspection
2. Assay Sheet of Calibrator SCS-1000
3. Printouts
4. Traceability & Uncertainty document