



**CALIBRATION CERTIFICATE**

User Name: M/s GLOBAL PATHOLOGY CENTRE  
C-16, Near DDU Hospital, Nanak Pura,  
Hari Nagar, New Delhi - 110064.

Equip. For Calibration	Digital Timer	Make / Model No.	BIO PLUS / -----
ULR No	CC205323000000493F	Range/ Size	0-24 hr.
Certificate No.	KRC/2023/BH/014-004	Least Count	0.01 hr.
Date Of Receipt Of UUC	11-Feb-2023	UUC Serial No.	----
Calibration Date	12-2-2023 to 14-2-2023	UUC I.D. No.	GPC/LAB/06
Suggested Due Date	11-Feb-2024	Location	Lab
Date Of Issue	14-Feb-2023	Visual Inspection	OK

Calibration Procedure	KRC/CPM/7.2-16-E-WI.		
Reference Standard	-----		
Discipline	ELECTROTECHNICAL (Time & Frequency)		
Environmental Condition	Temperature	25 ± 4 °C	Relative Humidity 50 ± 10 % RH

**Standard Equipment Used for Calibration**

Instrument Name	Make	LD No./Sr.No.	Traceability	Certificate No.	Suggested Due Date
DIGITAL TIMER	RED LION	---- / M4834A	GODREJ	C-220303-6-1	10-Mar-2023

**Calibration Results**

Sr. No.	Set Value on UUC (hr.)	Measured Value on Standard (second)	Expanded Uncertainty (second)
1	1	3600.351	±1.2
2	2	7200.73	±7.3
3	4	14400.20	
4	7	25200.50	±25
5	24	86400.30	

**Uncertainty of Measurement:** The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$  such that the coverage probability corresponds to approximately 95%.)

**Note:**

1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.
2. The results reported relate only to the above calibrated item.
3. This report should not be reproduced except in full without our prior permission in writing.
4. Calibration certificate without signature are not valid.
5. UUC stands for unit under calibration.

Calibrated By  
Calibration Engineer

**\*END OF REPORT\***

Page No.: 1 of 1

