




Trust & Care

**heartbeat**

Medical Systems

CALIBRATION REPORT

Report No: HBMS/CAL/22-23		Calibration date : 02.01.2023
		Calibration valid: 01.07.2023
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, UPGRADED PRIMARY HEALTH CARE CENTRE, THARAMANGALAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
I	Manufactured by	MICROSIC
ii	Model	T-10
iii	Serial No	UPHC/TM/CE/002

CALIBRATION WORKSHEET**LOCATION: LABORATORY****Tabulation For – Centrifuge**


S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	2	1.12
2	1200	1201	2	1.23
3	1500	1502	4	1.56
4	2000	1999	1	1.175

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

No. 23, 2nd Floor, jayasimmapuram,
Pappanaikenpalayam, Coimbatore – 37.

99400 28355, 97512 06106
88833 15662

hbmedicalsystem@gmail.com
hbmssales@gmail.com



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non - contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :