




CALIBRATION REPORT

Report No: HBMS/CAL/22-23		Calibration date: 02.01.2023
		Calibration due: 01.07.2023
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, UPGRADED PRIMARY HEALTH CARE CENTRE, THARAMANGALAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	THERMOSCIENTIFIC
II	Model	FINNPIPETTE F3
III	Value	100µl - 1000µl
IV	Serial No	RW11532


Details of the machine: machine calibrated and checked output.

Number of measurement	Channel I (Volume)		
	100µl	500µl	1000µl
1	100.04	500.09	999.98
2	100.05	500.07	1000.02
3	100.02	500.05	999.97
4	99.98	499.95	1000.02
5	100.03	500.08	1000.08
6	100.01	499.99	1000.04

For Heart Beat Medical Systems,



(Authorized signatory)



CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	Calibration date : 02.01.2023
	Calibration valid: 01.07.2023
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, UPGRADED PRIMARY HEALTH CARE CENTRE, THARAMANGALAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	FINNPIPETTE F3
II Model	10-100 µl
III Serial No	RW00632

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,


(Authorized signatory)




CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	Calibration date : 02.01.2023
	Calibration valid: 01.07.2023
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, UPGRADED PRIMARY HEALTH CARE CENTRE, THARAMANGALAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	FINNPIPETTE F3
II Model	10-100 µl
III Serial No	RW09146


Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.07	50.08	100.15
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.07
5	10.00	49.94	100.09
6	10.08	50.08	100.12

For Heart Beat Medical Systems,


(Authorized signatory)





CALIBRATION REPORT

Report No : HBMS/CAL/22-23		Calibration date : 02.01.2023
		Calibration valid: 01.07.2023
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, UPGRADED PRIMARY HEALTH CARE CENTRE, THARAMANGALAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
I	Manufactured by	ERBA
II	Model	BIOHIT
III	Value	200 µl
IV	Serial No	7012670

Details of the machine : machine calibrated and checked output.


Number of measurement	Channel 1 (Volume)
1	200.06
2	200.09
3	200.02
4	200.08
5	200.02
6	200.07

For Heart Beat Medical Systems,



(Authorized signatory)






CALIBRATION REPORT

Report No : HBMS/CAL/22-23		Calibration date : 02.01.2023
		Calibration valid: 01.07.2023
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, UPGRADED PRIMARY HEALTH CARE CENTRE,THARAMANGALAM ,SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
I	Manufactured by	ERBA
II	Model	BIOHIT
III	Value	20 µl
Iv	Serial No	7067918

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
1	20.06
2	20.09
3	20.02
4	20.08
5	20.02
6	20.07

For Heart Beat Medical Systems,



(Authorized signatory)

