



Date: 10/01/2023

Certificate of Calibration

Customer Name: Ujala Cygnus Rainbow Hospital, Agra

Model: Fully Automated Biochemistry Analyzer EM-200

Serial No.: 2002715

Calibration done date: 10/01/2023

Next Calibration due Date: On or Before: 09/01/2024

Lab Incharge: Dr. Prachi Mehta

*This is to certify that above mentioned product has been verified of calibration
By*

Checking Lamp Voltage -11.79

Checking gain of Photo meter for all Filters-OK

Checking Cell blanks value-OK

Checking laundry section-OK

Checking movement of probe-OK

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of
QC*

Calibration at site performed by Pushpendra Singh Bhadauria



Pushpendra Singh Bhadauria
Sr. Service Engineer
Agra