

Ref: SNS/AS/23030 Date: 07/04/2023 Tel No: +91 8584852151 E-Mail Id:sns4.kolkata@gmail.com

Calibration & Validation Certificate: 23030

Customer Name : Anubhav Life Care

2/8, Jessore Road (East)

Barasat

Kolkata - 700124

Instrument Model : Medonic M 32 Basic

Serial No. : 118008

Date of Calibration : 07/04/2023

Validity : 06/10/2023

This is to certify that the above mentioned product has been verified of calibration electronically and the instrument is working satisfactory.

Next Calibration : 07/10/2023

For SCIENCE & SERVICE

ANUPAM SAHA

Engineer



Company's GSTIN: 19ACMFS3034L1ZO, PAN No.: ACMFS3034L



Ref: SNS/AS/23030 Date: 07/04/2023 Tel No: +91 8584852151 E-Mail Id:sns4.kolkata@gmail.com

Calibration & Validation Certificate: 23030

Instrument Model : Medonic M 32 Basic

Serial No. : 118008

Calibrator : Boule Hematology Calibrator, Lot No. – 22301-34, Expiry – 21.04.2023

By comparing the data to the result of the standard counters in Medonic, the calibration for CBC – 6 parameters using the measurement standard material (Boule Calibrator) was completed. The calibration result of repeatability test is summarized in the following table. Please refer to the attached sheets for the details.

| SMP NO. | WBC | RBC | HGB | MCV | MPV | PLT |
|-------------------|-----|------|------|------|------|-----|
| 1 | 8.4 | 4.22 | 12.8 | 82.3 | 10.3 | 237 |
| 2 | 8.5 | 4.20 | 12.7 | 82.1 | 10.3 | 245 |
| 3 | 8.5 | 4.25 | 12.9 | 82.3 | 10.1 | 240 |
| 4 | 8.4 | 4.24 | 12.9 | 81.8 | 10.0 | 248 |
| 5 | 8.7 | 4.26 | 12.8 | 81.8 | 10.1 | 238 |
| Instrument Mean | 8.5 | 4.23 | 12.8 | 82.1 | 10.2 | 242 |
| Assay Mean | 8.4 | 4.23 | 12.7 | 82.8 | 10.0 | 235 |
| Tolerance Limit ± | 0.3 | 0.1 | 0.2 | 2 | 1.5 | 15 |

The Tolerance Limit represents the allowable limit for calibration, which was set up by Boule Medical and results are shown in attached sheet.

For SCIENCE & SERVICE

ANUPAM SAHA

Engineer



Company's GSTIN: 19ACMFS3034L1ZO, PAN No.: ACMFS3034L