



TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE
Precision is Our Destination.....

An NABL Accredited Calibration Laboratory
A Constituent Board of Quality Council of India

An ISO 9001 : 2015 Certified company

Calibration of Various Instruments & Testing Machines...



CC - 3125

Form No.: TMSCC/R/23

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE NO.:		2023/03/426	Page 1 of 2
ISSUE DATE :		17-Mar-2023	
ULR No. :		CC312523000002555F	
1.0	ISSUED TO:	M/s.: Anubhab Life Care 2/8, Jessore Road(East),Genjimill, Near Barasat Medical College,Barasat, North 24 pgs., W.B	

1.1	Service Request Form No. :	SRF/2023/03/11/01
1.2	Service Request Date :	11-03-2023
1.3	Location:	At Lab

Description identification of item to be calibrated:						
1.4	i	Name:	Micro Pipette	ii	Make:	Accupipette
	iii	Model / Type No.:	--	iv	S/L No.:	V90555
	v	I.D.No.:	ALC/MIP/V3	vi	Job Code No:	2023/03/426
	vii	Range:	(100 to 1000) µl	viii	Resolution:	2 µL
	ix	Accuracy	As Per ISO 8655-6	x	End User :	--

1.5	Full / Partial Calibration:	Full Calibration
-----	-----------------------------	------------------

1.6	Applicable specification of item to be calibrated: Accuracy / permissible limit:	Not Specified.
-----	--	----------------

1.7	Date of receipt of item:	11-Mar-2023	1.8	Date of calibration:	11-Mar-2023
-----	--------------------------	-------------	-----	----------------------	-------------

1.9	Calibration due on:	11-Mar-2024	2.0	Frequency of calibration once in:	366	Days
-----	---------------------	-------------	-----	-----------------------------------	-----	------

2.1	Environmental condition during calibration	Temperature:	25°C ± 4°C
		Humidity:	(50 ± 10)% RH

2.2	Basis of calibration:	SOP/04/02
-----	-----------------------	-----------

- The above Instrument has been calibrated over its range and the results are tabulated from page 2 of report.
- The Calibration Certificates relates only to the above DUC & reported results are valid at the time of and under the stated conditions of measurements.
- Errors if any, in this certificate shall be brought to notice within 45 days from the date of this certificate.
- NABL-133 guidelines are adopted for use of NABL symbol.
- This certificate shall not be reproduced, except in full, without the written permission of TMSCC
- Corrections/erasing invalidate the Calibration Certificate- exception to the 'Final Page or Part of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
- Laboratory Standards are traceable to National Standards
- In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.

2.3	Traceability : Standards used for calibration are traceable to National Standards through NABL accredited Laboratory.					
	Name of the Instrument	Sl. No./ Id No.	Certificate No.	Lab Certificate No.	Calibrated On	Due On
	Digital Weighing Balance	(TMSCC/EB/01) (Sl. No.- 14255716)	TSC/22-23/12101-25	CC-2231	15-11-2022	15-11-2023
	Digital Temperature Indicator with Sensor(RTD)	Sl. No.- 18K588073 (Sensor Sl. No.- 19102403)	TSC/22-23/12095-36		17-11-2022	17-11-2023



TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE
Precision is Our Destination.....

An NABL Accredited Calibration Laboratory
A Constituent Board of Quality Council of India
An ISO 9001 : 2015 Certified company

Calibration of Various Instruments & Testing Machines...



CC - 3125

Form No. : TMSCC/R/23						CC No. : 2023/03/426		
						Date : 17-03-2023		
						Page No. : 2 of 2		
CALIBRATION RESULTS								
Recorded Temperature : 20.1°C								
Sl. No.	Denomination Volume in μ l	Observed Reading at Ref. Std. Balance in g (Avg. of five readings)	Density of De-ionized Water in g/ml	Actual volume in μ l	Error in μ l	MPE in $\pm \mu$ l	Measurement Uncertainty in $\pm \mu$ l	Acceptance Criteria
1	100	0.09981	0.998183	99.99659	-0.0034	0.12	1.83	Pass
2	500	0.49907	0.998183	499.98297	-0.0170	0.20	1.83	Pass
3	1000	0.99815	0.998183	999.96594	-0.0341	0.20	1.83	Pass
Measurement Uncertainty at 95% confidence level where coverage factor, k=2								
Remarks: The above DUC has been calibrated over its above range & the readings observed are tabulated above. The reference standard used is traceable to National Standard.								
DUC: Device Under Calibration								
Physical status of the Instrument : Ok								

Calibrated By :

J. Bhattacharjee

(Sr. Calibration Engineer)

Checked By :

S. Chowdhury

(Quality Manager)

Approved By :

C. Ghosh

(Technical Director)

END OF CERTIFICATE

