



TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE
Precision is Our Destination.....

An NABL Accredited Calibration Laboratory
A Constituent Board of Quality Council of India

An ISO 9001 : 2015 Certified company
Calibration of Various Instruments & Testing Machines...



CC - 3125

Form No.: TMSCC/R/23

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE NO.:		2023/03/418		Page 1 of 2							
ISSUE DATE :		17-Mar-2023									
ULR No. :		CC312523000002547P									
1.0 ISSUED TO :		M/s.: Anubhab Life care 2/8, Jessore Road(East), Genjimill, Near Barasat Medical College, Barasat, North 24 pgs., W.B.									
1.1 Service Request Form No.:		SRF/2023/03/11/01									
1.2 Service Request Date :		11-03-2023									
1.3 Location :		At Lab									
1.4 Description identification of item to be calibrated :											
i Name :		Digital Thermometer		ii Make :		ACETEQ					
iii Model / Type No.:		ST 9269		iv S/L No.:		--					
v I.D.No.:		ALC/DIGTH/01		vi Job Code No.:		2023/03/418					
vii Range :		(-)50 to 300°C		viii Resolution :		0.1°C					
ix Accuracy :		As Per DKD R5 - 1		x End User :		--					
1.5 Full / Partial Calibration :		Partial Calibration									
1.6 Applicable specification of item to be calibrated: Accuracy / permissible limit :		Not Specified.									
1.7 Date of receipt of item :		11-Mar-2023		1.8 Date of calibration :		11-Mar-2023					
1.9 Calibration due on :		11-Mar-2024		2.0 Frequency of calibration once in :		366 days					
2.1 Environmental condition during calibration :		Temperature : 25°C ± 5°C									
		Humidity : (50 ± 20)% RH									
2.2 Basis of calibration :		SOP/10/02									
1. The above Instrument has been calibrated over its range and the results are tabulated from page 2 of report.											
2. The Calibration Certificates relates only to the above DUC & reported results are valid at the time of and under the stated conditions of measurements.											
3. Errors if any, in this certificate shall be brought to notice within 45 days from the date of this certificate.											
4. NABL-133 guidelines are adopted for use of NABL symbol.											
5. This certificate shall not be reproduced, except in full, without the written permission of TMSCC											
6. Corrections/erasing. invalidate the Calibration Certificate- exception to the 'Final Page or Part of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).											
7. Laboratory Standards are traceable to National Standards											
8. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '!' indicates no specification limit furnished. Acceptance criteria as per IS											
2.3 Traceability : Standards used for calibration are traceable to National Standards through NABL accredited Laboratory.											
Name of Instrument		Sl. No. / ID no.		Certificate No.		Lab Certificate No.		Calibrated On		Due On	
Digital Temperature Indicator with Sensor(RTD)		Sl. No.- 18K588073 (Sensor Sl. No.- 19102403)		TSC/22-23/12095-36		CC-2231		17-11-2022		17-11-2023	



TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE
Precision is Our Destination.....



An NABL Accredited Calibration Laboratory
A Constituent Board of Quality Council of India

An ISO 9001 : 2015 Certified company

Calibration of Various Instruments & Testing Machines...

CC - 3125

Form No.: TMSCC/R/23				CC No.:	2023/03/418
				Date :	17-Mar-2023
				Page No.:	2 of 2
CALIBRATION RESULTS					
Sl. No.	Denomination set in °C	Observed Reading at Ref. Std. in °C (Avg. of five readings)	Error in °C	Uncertainty in ± °C	Acceptance Criteria
1	-30	-29.979	0.021	0.18	Pass
2	-20	-19.986	0.014	0.18	Pass
3	-10	-9.993	0.007	0.18	Pass
4	-5	-4.996	0.004	0.18	Pass
5	5	4.996	-0.004	0.18	Pass
6	8	7.994	-0.006	0.18	Pass
7	20	19.986	-0.014	0.18	Pass
8	50	49.964	-0.036	0.33	Pass
9	100	99.928	-0.072	0.33	Pass
10	250	249.821	-0.179	0.33	Pass
Maximum Permissible Error :		± 0.3% of rdg			
Measurement Uncertainty at 95% confidence level where coverage factor, k=2					
REMARKS : The DUC has been calibrated over its range. The readings observed are tabulated above. The reference standard is traceable to National standard.					
DUC: Device Under Calibration.					
Physical Status of the DUC : OK					

Calibrated By :


R. Ghosh

(Calibration Engineer)

Checked By :


C. Ghosh

(Technical Director)

Approved By :



P. K. Modak
(Sr. Calibration Engineer)

END OF CERTIFICATE

