



Scientific Temper



Honesty, Humility



NORTHLAB (INDIA) PVT. LTD.

Laboratory : Plot No. 1, S.V Nagar, Perumalpatthu, Thiruvallur - 602 024.
Ph.: +91 89399 23940 / 89399 40050 E-mail: sales@northlab.in Website: www.northlab.in



CC-2457

NORTHLAB

Calibration Certificate

Cert. No. : 2360622-T211 ULR NO : CC245722900002542F

Date of Issue : 23-Jun-2022 Page : 1 of 1

Customer : SAICTC

Date Received : 22-Jun-2022

Date Calibrated : 22-Jun-2022

Due Date : 21-Jun-2023

Ambient Temp. : 26.3±3°C

Relative Humidity : 61.4±15% R.H

Address : Govt General Hospital,
Kamarajar Salahi,
Karnikal - 609 602.

Instrument : REFRIGERATOR

Manufacturer : HAIER

Model No. : HRD-2623/2020

Part No. : --

Location : SAICTC

Serial No. : 170/1

Instrument ID. : --

Range (Res.) : --

Accuracy / MPE : --

Test Results : As Found

METHOD OF CALIBRATION

1. The instrument was calibrated as per method no. TEC at customer premises in accordance with ISO/IEC 17025:2017(E).

Master Instrument Used

1. Temperature Sensor (PT-100)

2. Multifunction Process Calibrator

Cert. No.

T021422

Cal. Date

04-Feb-2022

Validity

03-Feb-2023

Traceability

NLI


NLI

RESULTS OF THERMAL CALIBRATION

INSTRUMENT	MASTER	TOLERANCE	STATUS	MEASUREMENT
SETTING(°C)	READING(°C)	(±)°C		UNCERTAINTY(±)°C
TOP	-15.6	--	--	0.59
MIDDLE	2.1	--	--	0.59
BOTTOM	2.7	--	--	0.59

REMARKS / DEVIATIONS IF ANY

(SEE BACK OF THIS PAGE)


S. Gunasekaran
Calibration Officer



A. Murugappan / L. Muthuraman / S. Poetha
Authorised Signatory
NorthLab (I) Pvt. Ltd

X:\Cert-2022\NABL-CertTemp-Saictc-1\COMMON HEALTH\2360622-T211\196E.XLS\196E1

End of Certificate



“காவினால் ஆகிய ஆக்கம் அளவிறந்து
ஆவது போலக் கெடும்” -குறள் 283

The property, which is acquired by fraud, will pass all bounds
and perish, even while it seems to increase. - Kural 283