

## AAMXBIOMEDICALS

Shop No. 01 Shiv Mandir Tara Nagar District Sonepat-131001, State - Haryana, India

Email: <u>aamxbiomedicals@rediffmail.com</u>, Mob: +91 9813151502

## **Calibration Certificate**

Following voltages have been checked & adjusted on Bio Chemistry Analyzer Trace 200 (Marketed By AAMX Biomedicals) Sr. No. AES2K2KH04065 at Elixir Diagnostic Centre, A-5/238, Ground Floor, Near A-5 DDA Market, Paschim Vihar, New Delhi-110063.

System Voltages DC are in the specified range -Checked by Digital Multimeter are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Main Power Supply	219 VAC	220 +/- 5.0 VAC
2	24 Voltage Supply	24.2 VDC	24.0 +/- 0.5 VDC
3	12 Voltage Supply	12.2 VDC	12.0 +/- 0.5 VDC
4	Liquid Level Sensor Board Supply	4.52 VDC	4.50 +/- 0.1 VDC

System Temperature values are in the specified range –Checked by Digital Thermometer are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Reaction Disk Temp.	37 °C	37 +/- 0.4 °C
2 .	Reagent Compartment Temp.	10 °C	8 to 12 °C

Photometer Adjustment Values are within range as recommended:

S.No.	Module	Observed Value	Acceptable Range
1	Background Count for all Wavelength	32,000 to 49,000	20,000 to 62,000

Washing Dispense are Ok volume dispensed measured on Digital Weighing Machine:

S.No.	Module	Observed Value	Acceptable Range
1	Interior Washing	3.01 mL	3.00 +/- 500uL
2	Exterior Washing	3.01 mL	3.00 +/- 500uL

- > All other mechanisms have been checked through software and found OK.
- > Performed calibration for End Point, Kinetic & Fixed Time-Kinetic. Run control & sample, Results are satisfactory.
- ➤ Next Calibration will be performed on 01/05/2023.

For AAMX Biomedicals,



Date: 02/05/2022