

CALIBRATION REPORT

To,
Kamal Diagnostics,
Chinchwad, Pune.

Date- 07/03/2023

Dear sir/Madam,

As desired by you please find enclosed herewith the report.

Name of System	EasyElectrolytes Analyzer
Sr. No.	0122041007
Calibration done on	06 th March 2023
Next calibration Due on	05 th March 2024
Calibrated By	Mr. Abhijit Jagtap


Instrument calibration Values,

	Na	k	cl
Results	145.3	3.92	98.6
Mean	142.9	3.9	98.83
Range	127.54-158.26	3.57-4.23	84.13-113.53
Cal. values	58.1	58.3	50.2

All value is within range.

Checked instrument all found ok.

With Regards



Crystal Diagnostics
Service Engineer.