



NENO TECHNICAL SERVICES

(Near to True Value)



CALIBRATION CERTIFICATE

Certificate No. : NTS/B/S-8708

User Name : REGIONAL RESEARCH INSTITUTE OF UNANI MEDICINE

Chandbali By-pass Road, Near Rural Police Station, Bhadrak-756100, Odisha

Gauge For Calibration	REFRIGERATOR	
Make	Godrej	
Range	2 to 10°C	
Least Count	1°C	
Instrument Id. No.	RRIUM/REF-01	
Instrument Sr. No.	-	
Location	PATHOLOGY	
Visual Inspection	OK	
Material Receipt Date		ON SITE
Date of Calibration		17.10.2022
Suggested Due Date of Calibration		16.10.2023
ENVIRONMENTAL CONDITION		
Temperature	30 ± 10 °C	
Relative Humidity	50 ± 10 %RH	

Equipment & Master Used For calibration	
01. Temp. Indicator with RTD Sensor	
Make	: SMK
Sr. No.	: 02120034
Certificate No.	: PMM/14132/1
Next Due Date	: 22.11.2022
Traceability	: PMMPL

OBSERVATION RESULTS

SL. NO.	UUC VALUE (°C)	MASTER VALUE (°C)	± Expanded Uncertainty At 95% Confidence Level (k=2) (°C)
1.	2	1.9	± 0.9
2.	3	2.9	± 0.9
3.	5	4.9	± 0.9
4.	8	7.8	± 0.9
5.	10	9.8	± 0.9

Calibration Procedure Based On : CP/T-04

NOTE :-

1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid.
4. All our certificates are Traceable to National standard.
5. UUC : Unit under calibration, GUT : Gauge under testing.

Calibrated By (D. Rout)

F03(5.10/01)

Page 1 of 1

Issued / Approved By

(D. Majhi)

Dy.T.M. / Quality Manager

ISO/IEC: 17025:2017 Accredited Lab

Plot No. : MIG-154, Baramunda Housing Board Colony, Bhubaneswar - 751003, Khurda, Odisha
Ph. : 0674-2355490, Mob. : 7894892390, 9437568822, 9438268493, E-mail : nenoindia5@gmail.com, Web : odisha@neno.org.in

Head Office : Neno Technical Services, A-2, Sector-92, Noida - 201304