



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - OCTOBER 2022



PC-1024

Lab Name: Govt.UGPHC, Thondi

Lab No: 7472

Constituent Group: Chemistry III

Date of Result Entered: 19/10/2022

PT Item: Lyophilized Serum

Date of Report Published: 03/11/2022

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	SI	Your Value	SI	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1837	346.52	7.17	24.84	385 mg/dL	1.55	1.16
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1831	1.36	10.34	0.14	1.5 mg/dL	0.99	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	109.51	6.27	6.87	108 mg/dL	-0.22	0.33
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt.UGPHC, Thondi
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, RAMANATHAPURAM DISTRICT.
 Thondi
 TAMILNADU623409

Coordinator Contact Details:
 Email: clnqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - NOVEMBER 2022



PC-1024

Lab Name Govt.UGPHC, Thondi

Lab No

7472

Constituent Group Chemistry III

Date of Result Entered :

20/11/2022

PT item Lyophilized Serum

Date of Report Published :

06/12/2022

Slit No Analyte Method Principle Name

Analyzer Name

No of Participants

DV

Participants SD CV

Your Value

SPC

U

Slit No	Analyte	Method	Principle Name	Analyzer Name	No of Participants	DV	Participants SD	CV	Your Value	SPC	U	
1	GLUCOSE III	GOD-POD III		Robonik (Semi automatic / Fully Automatic)	1868		82.95	8.62	7.15	80 mg/dL	-0.41	0.33
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III		Robonik (Semi automatic / Fully Automatic)	1843		6.98	7.13	0.50	7.1 mg/dL	0.24	0.02
3	CHOLESTROL III	CHOD-PAP III		Robonik (Semi automatic / Fully Automatic)	1767		112.60	5.74	6.46	106 mg/dL	-1.02	0.31
SDI Range					Interpretation							
Within -1.00 to +1.00					Excellent.							
Within ±1.01 to ±2.00					Good.							
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.							
Beyond ±3.0					Unacceptable performance. Action Signal.							

LAB ADDRESS :

Govt.UGPHC, Thondi
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, RAMANATHAPURAM DISTRICT.
Thondi
TAMILNADU623409

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components





CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name Govt.UGPHC, Thondi

Lab No 7472

Constituent Group Chemistry III

Date of Result Entered : 20/12/2022

PT item Lyophilized Serum

Date of Report Published : 05/01/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Partic. CV	SD	Your Value	BI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1844	239.20	6.25	14.96	219 mg/dL	-1.35	0.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.91	8.24	0.32	4.2 mg/dL	0.90	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1766	114.94	5.41	6.21	112 mg/dL	-0.47	0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt.UGPHC, Thondi

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, RAMANATHAPURAM DISTRICT.

Thondi

TAMILNADU623409

Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components



VIEW LAB MONTHLY SUMMARY

Lab Name Govt.UGPHC, Thondi
 Lab No 7472
 Month January
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/01/2023

Date of Report Published : 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	118 mg/dL	-0.24	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	2.2 mg/dL	-0.05	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	110 mg/dL	-0.32	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Lab Name: Govt.UGPHC, Thondi
 Lab No: 7472
 Constituent Group: Chemistry III
 Chemistry III
 PT Item: Lyophilized Serum
 Date of Result Entered: 21/02/2023
 Date of Report Published: 08/03/2023

Sl No	Analyte	Method / Proprietary Name	Analyzer Name	No of participants	CV	Participants	CV	Participants	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	164 mg/dL	-0.87	0.47	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.2 mg/dL	-0.65	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	104 mg/dL	-1.49	0.28	
SDI Range Interpretation											
Within -1.00 to +1.00 Excellent.											
Within ±1.01 to ±2.00 Good.											
Within ±2.01 to ±2.99 Accept with caution. Warning Signal.											
Beyond ±3.0 Unacceptable performance. Action Signal.											

LAB ADDRESS :
 Govt.UGPHC, Thondi
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, RAMANATHAPURAM DISTRICT.
 Thondi
 TAMILNADU623409

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Verified By
Mp
 20/3/2023
 Dist. Microbiologist



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name: Govt UGPHC, Thondi
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 7472
Date of Result Entered: 20/03/2023
Date of Report Published: 04/04/2023

Name: Analyte: Method: Principle Name: Analyzer Name: No. of Participants: CV: SD: Mean Value: SD: U

No	Analyte	Method	Principle Name	Analyzer Name	No. of Participants	CV	SD	Mean Value	SD	U		
1	GLUCOSE III	GOD-POD III		Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	88 mg/dL	6.82	-0.51	0.31	
2	CREATININE III	Jaffes Kinetic - Alkaline picrate III		Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	1.3 mg/dL	0.12	-0.25	0.01	
3	CHOLESTROL III	CHOD-PAP III		Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	5.30	113 mg/dL	0.36	0.23
SDI Range					Interpretation							
Within -1.00 to +1.00					Excellent.							
Within +1.01 to +2.00					Good.							
Within +2.01 to +2.99					Accept with caution. Warning Signal.							
Beyond +3.0					Unacceptable performance. Action Signal.							

LAB ADDRESS :
Govt UGPHC, Thondi
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, RAMANATHAPURAM DISTRICT.
Thondi
TAMILNADU623409

Coordinator Contact Details:
Email: clingc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore