

Ph. : (O) + 91 265 3927448/49/50  
2632148/49/50  
Fax : (O) + 91 265 2632151

## LABLINE EQUIPMENTS PVT. LTD.

An ISO 9001 : 2008 & 13485: 2003 Accredited Company



Mfgr. Of Medical Equipments

614/1, G.I.D.C. B/H Himalaya Machinery, Makarpura, Vadodara - 390 010, Gujarat, INDIA  
E-mail : haresh@lablineindia.com • mail@lablineindia.com • Web Site: www.lablineindia.com

### CALIBRATION CERTIFICATE

**CUSTOMER :**

LIFETECH DIAGNOSTICS  
234/3, Sri Nagar Colony,  
Narasothipatti, Kuranguchavadi,  
Salem - 636 004

CERTIFICATE NO. **LL/257/20-21**

**DEALER :**

Labtech Medico (P) Ltd  
XI-434 Chakkiath Joseph Memorial Building,  
Parappuram Junction, Karukutty P.O,  
Ernakulam District, Kerala-683576.

DESCRIPTION OF ITEM : Digital Medical Centrifuge ( CF 655 ) with Rotor RT 07 (16 x 15 ml Swing out Head)  
CALIBRATION DATE : 08/08/2022

**MODEL NO.: CF 655 M/C.Sr.No. 594 - A**

**RANGE : 1000 - 3500 RPM**

DETAILS OF EQUIPMENTS USED FOR CALIBRATION :--

SR. NO.	DESCRIPTION	CERTIFIED BY	CERTIFICATE NO.	VALIDITY
1	DIGITAL TACHOMETER	G.I.I.	I - 3776	1 YEAR

OBSERVATION :

R.P.M. CALIBRATION \* [ For Centrifuges only ]

POINT	INSTRUMENT R.P.M. DISPLAYED	CALIBRATED TACHO METER READING	ERROR OBSERVED	ACCEPTED ERROR	RESULT
1	1000	1051	-51	+/- 100	OK
2	2500	2537	-37	+/- 100	OK
3	3500	3542	-42	+/- 100	OK

TEMPERATURE CALIBRATION \* : [For Dry bath incubators only]

POINT	INSTRUMENT TEMPERATURE DISPLAYED (°C)	CALIBRATED THERMOMETER TEMP. (°C)	ERROR OBSERVED	ACCEPTED ERROR	RESULT

NEXT CALIBRATION DUE DATE : 07/08/2023

Labline Equipments Pvt. Ltd.

APPROVED BY:  Director

Labline Equipments Pvt. Ltd.  
614/1, G.I.D.C. Estate,  
Makarpura, Vadodara - 390 010.  
Phone: 3927448, 49, 50.

INSPECTION DEPARTMENT

\* : RPM & temperature readings are taken on "LABLINE EQUIPMENTS PVT. LTD." standard testing panels and instruments. All readings are taken under ideal test condition.

\*\* : As required by the user.