



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name **SRI ARUNA COMPUTER DIAGNOSTIC CENTER** Lab No **13169**  
 Constituent Group **Chemistry II** Date of Result Entered : **19/12/2022**  
 PT item **Lyophilized human serum based** Date of Report Published : **05/01/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1249	163.25	9.53	15.55	150 mg/dL	-0.85	0.88
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation)	808	86.08	13.67	11.77	90 mg/dL	0.33	0.83
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	648	1.10	13.68	0.15	1.1 mg/dL	0.00	0.01
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1024	2.16	17.66	0.38	3 mg/dL	2.20	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	928	4.57	15.28	0.70	4.8 g/dL	0.33	0.05
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	919	2.73	14.20	0.39	2.4 g/dL	-0.85	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1059	5.50	15.68	0.86	5.7 mg/dL	0.23	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1228	94.08	15.32	14.41	93 mg/dL	-0.07	0.82
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1042	94.26	18.90	17.82	100 mg/dL	0.32	1.10

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

**SRI ARUNA COMPUTER DIAGNOSTIC CENTER**  
**TRICHY MAIN ROAD, VEPPUR**  
**CUDDALORE**  
**TAMILNADU606304**

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***