

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	<b>SRI ARUNA COMPUTER DIAGNOSTIC CENTER</b>	<a href="#">Details About Robust Analysis</a>								
<b>Lab No</b>	<b>13169</b>	<a href="#">Detail About Monthly Summary</a>								
<b>Month</b>	<b>March</b>	<a href="#">Detail about SDI</a>								
<b>Year</b>	<b>2023</b>									
<b>Constituent Group</b>	<b>Chemistry II</b>									
<a href="#">Click on the analyte to view Graphical Data</a>		<input type="button" value="All Analyser Result"/> <input type="button" value="Print"/>								
Date of Result Entered : 20/03/2023										
Date of Report Published : 04/04/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	<a href="#">GLUCOSE II</a>	GOD-POD II	Any Analyser (Automation / Semi Automation)	1561	308.70	9.98	30.81	300 mg/dL	-0.28	1.56
2	<a href="#">UREA II</a>	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation)	945	56.76	13.22	7.50	64 mg/dL	0.96	0.49
3	<a href="#">CREATININE II</a>	Jaffes End point II	Any Analyser (Automation / Semi Automation)	769	1.84	13.24	0.24	2 mg/dL	0.66	0.02
4	<a href="#">T.BILIRUBIN II</a>	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1367	5.35	15.48	0.83	6 mg/dL	0.78	0.04
5	<a href="#">T-PROTEIN II</a>	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1230	5.53	10.94	0.60	5.5 g/dL	-0.05	0.03
6	<a href="#">ALBUMIN II</a>	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1198	3.31	9.84	0.33	3 g/dL	-0.95	0.02
7	<a href="#">URIC ACID II</a>	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1224	4.62	19.05	0.88	5 mg/dL	0.43	0.05
8	<a href="#">CHOLESTROL II</a>	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1526	112.67	12.38	13.95	137 mg/dL	1.74	0.71
9	<a href="#">TRIGLYCERIDE II</a>	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1378	140.75	14.47	20.37	158 mg/dL	0.85	1.10
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>