



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO. : 221125A/N-105 Page : 1 of 1

CALIBRATED ON. : 25 / 11 / 2022 Ref. Std. : ISO:8655:6  
CALIBRATION DUE. : 24 / 05 / 2023 C.P. No. : ACS/CP/08

SRF No. & DATE : 221125A/N Dt. 25.11.2022 Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC342722000001031F Dt. 28.11.2022 Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.  
Shop No-203, 2<sup>nd</sup> Floor Landmark Complex,  
Prahlad Nagar 100ft Road, Near Seema Hall, Ahmedabad

ITEM CALIBRATED : Micro Pipette  
Make : Thermo Scientific Range : (100 to 1000) µl  
L. Count : ID No. : RCL/BIO/PIP-05  
Location : Clinical Biochemistry S. No. : SH81961

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		100	100.35	(-) 0.35
1000		500	501.61	(-) 1.61
		1000	1003.22	(-) 3.22

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.12 µl at coverage factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

