



Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC-2645

CALIBRATION CERTIFICATE


AVI/certificate/F-12D		Calibration Certificate for Digital Thermometer			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on	Date of Issue:	ULR No.:	CC26452200000469F	
14.12.2022		13.12.2023	15.12.2022	Certificate No.:	AVI2212029	
SRF No.	2200241	SRF Date	SRF No.	Job No	2212029	Job date
						14.12.2022
Discipline : Thermal calibration			Group : Temperature			
Customer's Name & Address:			Disha CRL Laboratory 74, Sopariwala baug Building, J.M Street, Opp.KEM Hospital, Parel, Mumbai-400012			
Customer Reference/Challan No			-			
Receipt Date			14.12.2022			
Condition on receipt			Satisfactory			
Calibration Carried out:			Onsite			
Description of item Calibrated:						
Equipment			Multi- Thermometer			
Make			Sigma			
Model			-			
Serial No.			1908010169			
Identification No.			DL/TM/01			
Range			(-50 to 300) °C			
Working			(-50 to 300) °C			
Least Count/Res			0.1 °C			
Location			-			
Accuracy			-			
Work Instruction No.			AVI/WI/14			
Environmental Condition			Temp. (°C): (25±5) Relative Humidity (% RH) (30 to 75)			
Standard(s) used for Calibration:			Digital Indicator with Temperature Sensor (4 wire), Sr. No. O19/420, 1677, Due Date 29/08/2023. Cal Certificate No.- M-220827-10-4, Cal By Godrej & Boyce, CC-2634			

OBSERVATION TABLE

Sl. No.	Range °C	Standard °C	UUC Reading °C	Error °C	Expanded Uncertainty (At 95% CL, K=2), (±°C)
1	(-50 to 300) °C	-39.81	-40.0	-0.19	0.47
2		10.23	10.0	-0.23	
3		50.28	50.0	-0.28	

Calibrated By

(Prathamesh Mestri)
Calibration Engineer



Authorised Signatory

Janardan Chavan
(Technical Manager)

----- End Of Certificate -----

A-221, Amargian Indl. Complex, Opp. New S. T. Stand, Khopat, Thane (w)-400601. (Mumbai)

Tel. No.022-25477638/40 *Email:calibration@avisscientific.com,md@avisscientific.com www.avisscientificindia.com



Avi Scientific (India)

CALIBRATION LABORATORY

CALIBRATION CERTIFICATE


AVI/certificate/F-12D		Calibration Certificate for Digital Thermometer			Page No: 01 of 01	
Date of Calibration:	Next Cal. due on:	Date of Issue:		Certificate No.: AVI 2212030		
14.12.2022	13.12.2023	15.12.2022				
SRF No.	2200241	SRF Date	SRF No.	Job No.	2212030	Job Date 14.12.2022
Discipline : Thermal calibration			Group : Temperature			
Customer's Name & Address:			Disha CRL Laboratory 74, Sopariwala baug Building, J.M Street, Opp.KEM Hospital, Parel, Mumbai-400012			
Customer Reference/Challan No			-			
Receipt Date			14.12.2022			
Condition on receipt			Satisfactory			
Calibration Carried out:			Onsite			
Description of item Calibrated:						
Equipment			Multi- Thermometer			
Make			Sigma			
Model			-			
Serial No.			1908010169			
Identification No.			DL/TM/01			
Range			(-50 to 300) °C			
Working			(-50 to 300) °C			
Least Count/Res			0.1 °C			
Location			-			
Accuracy			-			
Work Instruction No.			AVI/WI/14			
Environmental Condition			Temp. (°C):		(25±5)	
			Relative Humidity (% RH)		(30 to 75)	
Standard(s) used for Calibration:			Digital Indicator with Temperature Sensor (4 wire), Sr. No. O19/420, 1677, Due Date 29/08/2023. Cal Certificate No.- M-220827-10-4, Cal By Godrej & Boyce, CC-2634			

OBSERVATION TABLE

Sl. No.	Range °C	Standard °C	UUC Reading °C	Error °C	Expanded Uncertainty (At 95% CL, K=2),(±°C)
1	(-50 to 300) °C	60.31	60.0	-0.31	0.47
2		75.37	75.0	-0.37	
3		100.41	100.0	-0.41	
4		250.46	250.0	-0.46	

Calibrated By

(Prathamesh Mestri)
Calibration Engineer



Authorised Signatory

(Janardan Chavan)
(Technical Manager)

----- End Of Certificate -----

A-221, Amargian Indl. Complex, Opp. New S. T. Stand, Khopat, Thane (w)-400601. (Mumbai)

Tel. No.022-25477638/40 *Email:calibration@avisscientific.com,md@avisscientific.com www.avisscientificindia.com