




Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC-2645






CALIBRATION CERTIFICATE

AVI/certificate/F-12D		Calibration Certificate for Temperature Sensor With Indicator			Page No 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue:	ULR No.:	CC264522000000459F	
14.12.2022		13.12.2023	15.12.2022	Certificate No.:	AVI2212022	
SRF No.	2200241	SRF Date	14.12.2022	Job No	2212022	Job date 14.12.2022
Discipline : Thermal			calibration Group : Temperature			
Customer's Name & Address:			Disha CRL Laboratory 74, Sopariwala baug Building, J.M Street, Opp.KEM Hospital, Parel, Mumbai-400012			
Customer Reference/Challan No			-			
Receipt Date			14 12 2022			
Condition on receipt			Satisfactory			
Calibration Carried out:			ONSITE			
Description of item Calibrated:						
Equipment			Freezer			
Make			Western			
Model			SRC 380-GL			
Serial No.			51632210402535			
Identification No.			DL/FR/02			
Range			(2 to 8)°C			
Working Range			(2 to 8)°C			
Least Count/Res			-			
Location			Lab			
Work Instruction No.			AVI/WI/14			
Environmental Condition			Temp (°C) (25±5) Relative Humidity (% RH) (30 - 75)			
Standard(s) used for Calibration:			Digital Indicator with Temperature Sensor (4 wire), Sr No O19/420, 1677, Due Date 29/08/2023. Cal Certificate No - -M-220827-10-4, Cal By, Godrej & Boyce Mfg Co Ltd., CC-2634			

OBSERVATION TABLE

Range °C	Location	Standard Reading	UUC Reading	Error	Expanded Uncertainty (At 95% CL, K=2), °C
		°C	°C	°C	
(2 to 8)°C	Freezer	4.67	5	0.33	0.47

Note: calibration point is taken as per customer request

Calibration Engineer  (Niranjan Rajguru) Calibration Engineer		Authorised Signatory  (Janardan Chavan) Technical Manager
--	--	---

-----End of the certificate-----