



Analog Healthcare Private Limited

CALIBRATION CERTIFICATE

Customer Name	DISHA CRL		
Department	Hematology		
Address	Mumbai		
Instrument	Mindray BC-5150 ANALYZER	Date of Calibration	12.11.2023
Serial No	SR-13004984	Next Calibration Date	12.11.2024
		Calibrator Lot Number	BC 1709B
		Calibrator Expiry date	12.03.2024

WB calibration data

	WBC	RBC	HGB	MCV	PLT
Target	9.04	4.57	13.5	89.4	254
WB 1#	8.60	4.06	12.6	84.4	233
WB 2#	8.72	4.08	12.5	84.2	240
WB 3#	8.71	4.09	12.6	83.9	244
WB 4#	8.60	4.08	12.6	84.5	247
WB 5#	8.74	4.09	12.6	83.9	245
WB 6#	8.77	4.10	12.5	84.3	242
WB 7#	8.57	4.11	12.7	84.2	241
WB 8#	8.55	4.07	12.5	84.0	241
WB 9#	8.51	4.05	12.6	83.7	234
WB 10#	8.69	4.07	12.6	83.8	238
Mean	8.65	4.08	12.6	84.1	241
CV%	1.0	0.5	0.4	0.3	1.9
Old Factor (%)	100.00	105.00	102.00	110.00	98.00
New Factor (%)	100.00	106.00	98.00	97.00	100.00

This is to certify that the Mindray BC 5150 Hematology Analyzer Sr.No: **SR-13004984** system, installed at **DISHA CRL**, Mumbai has been calibrated using the standard reference material CBC-SDMR.



Abhishek Reshamwala
Sr. Service Engineer
For Analog Healthcare Pvt. Ltd