

	Form Title	Calibration of Refrigerated Centrifuge	Annexure II
	Relevant SOP No.:	BA040-00	Page No.: 1 of 1

Details of Instrument used for Calibration

Calibrator	Instrument ID	Next Calibration Due Date	Certificate No.
Tachometer	BA/DT/049	07 DEC 2023	22120707000109
Temperature Probe	CR/TM/025	24 MAY 2023	22052404001509
Digital Stopwatch	CR/DSW/013	24 MAY 2023	SL2205EL0171-001

Instrument Details of Refrigerated Centrifuge

Instrument ID:	BA/CEF/029		
Calibration Date:	14 DEC 2022	Next Calibration Due Date:	30 MAY 2023

Sr. No.	Parameter	Standard Values	Observed Reading	Acceptance Criteria
1	Speed (rpm)	1000	1010	± 50 of set speed
		3000	2990	
		4000	3995	
2	Time (Min)	5	5	± 15 sec of set value
		15	15	
		30	30	
3	Temperature (°C)	4	4	± 2.0°C
		10	11	
		25	24	

Results: Meets acceptance criteria / Does not meet acceptance criteria

Remarks: —

Calibrated by:
(Signature & date)

Ch. Rajitha
Christina
14 DEC 2022

Checked by:
(Signature & date)

SPM
14 DEC 2022