

Cert. No.	: 2360622-E114	ULR NO	: CC245722600003796F
Date of Issue	: 23-Jun-2022	Page	: 1 of 1
Customer	: SAICTC	Date Received	: 22-Jun-2022
Address	: Govt General Hospital, Kamarajar Salai, Karaikal - 609 602.	Date Calibrated	: 22-Jun-2022
		Due Date	: 21-Jun-2023
		Ambient Temp.	: 25±2°C
		Relative Humidity	: 60±10% r.h.
Instrument	: <u>TIMER (CENTRIFUGE)</u>	Serial No.	: CKC237
Manufacturer	: ELTEK	Instrument ID.	: ---
Model No.	: TC 650D	Range (Res.)	: 0-99min
Part No.	: ---	Accuracy / MPE	: ---
Location	: SAICTC	Test Results	: As Found

**METHOD OF CALIBRATION**

1. The instrument was calibrated as per method no. ETIM at customer premises in accordance with ISO/IEC 17025:2017(E).

<u>Master Instrument Used</u>	<u>Cert. No.</u>	<u>Cal. Date</u>	<u>Validity</u>	<u>Traceability</u>
1. Digital Timer	TSC/21-22/13349-1	25-Dec-2021	25-Dec-2022	TRANSCAL

**RESULTS OF ELECTRO TECHNICAL CALIBRATION**

MASTER VALUE				DUC READING				CORRECTION		TOLERANCE (±) s	STATUS
Hr.	min.	sec.	1/100 s.	Hr.	min.	sec.	1/100 s.	sec.	1/100 s.		
00	05	00	01	---	05	00	---	00	01	---	---
00	10	00	04	---	10	00	---	00	04	---	---
00	15	00	06	---	15	00	---	00	06	---	---
00	30	00	09	---	30	00	---	00	09	---	---
00	45	00	13	---	45	00	---	00	13	---	---

**REMARKS / DEVIATION IF ANY**

(SEE BACK OF THIS PAGE)

1. This result has an expanded uncertainty of ±0.47s with a coverage factor k=2 at approximately 95% confidence level.



*(Handwritten signature)*