



Scientific Temper



Honesty, Humility

**NORTHLAB (INDIA) PVT. LTD.**Laboratory : Plot No. 1, S.V Nagar, Perumalpattu, Thiruvallur - 602 024.
Ph.: +91 89399 23940 / 89399 40050 E-mail: sales@northlab.in Website: www.northlab.in**NORTHLAB****Calibration Certificate**

CC-2457

Cert. No. : 2870622-T03	ULR NO : CC245722400001301F
Date of Issue : 24-Jun-2022	Page : 1 of 1
Customer : THE MEDICAL OFFICER INCHARGE	Date Received : 23-Jun-2022
Address : Community Health Centre, T.V.Malai Road, Mannadipet, Puducherry - 605501.	Date Calibrated : 24-Jun-2022
	Due Date : 23-Jun-2023
	Ambient Temp. : 23±3°C
	Relative Humidity : 60±15% R.H
Instrument : <u>DIGITAL THERMOMETER WITH SENSOR</u>	Serial No. : 201506012118
Manufacturer : III INSTRUMENTICS	Instrument ID. : ---
Model No. : ---	Range (Res.) : -50-300°C(0.1°C)
Part No. : ---	Accuracy / MPE : ±1°C
Location : ---	Test Results : As Found

METHOD OF CALIBRATION

1. The instrument was calibrated as per method no. TTS at Northlab in accordance with ISO/IEC 17025:2017(E).

Master Instrument Used	Cert. No	Cal. Date	Validity	Traceability
1. Platinum Resistance Thermometer (PRT) With DMM	TSC/21-22/5622-2	03-Aug-2021	02-Aug-2022	TRANSCAL

RESULTS OF THERMAL CALIBRATION

MASTER READING (°C)	INSTRUMENT READING (°C)	CORRECTION (°C)	TOLERANCE (±) °C	STATUS	MEASUREMENT UNCERTAINTY (±) °C
-19.92	-19.7	-0.22	1.0	PASS	0.39
0.24	0.5	-0.26	1.0	PASS	0.39
50.28	50.0	0.28	1.0	PASS	0.39
150.05	149.7	0.35	1.0	PASS	0.39
249.97	249.5	0.47	1.0	PASS	0.39

REMARKS / DEVIATIONS IF ANY (SEE BACK OF THIS PAGE)

1. The instrument is within the Maximum Permissible Error.

D.Kamali
D.Kamali
Calibration Officer



A.Murugappan / L.Muthuraman / S.Picetha
A.Murugappan / L.Muthuraman / S.Picetha
Authorised Signatory
NorthLab (I) Pvt. Ltd

Y:\Cert-2022\NABL-Cert\Temperature\J-R\PONDICHERRY A1D5\2870622-T03\96E1.XLS

End of Certificate

“களவினால் ஆகிய ஆக்கம் அளவிறந்து
ஆவது போலக் கெடும்” -குறள் 283



The property, which is acquired by fraud, will pass all bounds
and perish, even while it seems to increase. - Kural 283