



Scientific Temper



Honesty, Humility



CC-2457

**NORTHLAB (INDIA) PVT. LTD.**Laboratory : Plot No. 1, S.V Nagar, Perumalpattu, Thiruvallur - 602 024.
Ph.: +91 89399 23940 / 89399 40050 E-mail: sales@northlab.in Website: www.northlab.in**Calibration Certificate**

Cert. No.	: 2360622-T208	ULR NO	: CC245722900002540F
Date of Issue	: 22-Jun-2022	Page	: 1 of 1
Customer	: THE MEDICAL OFFICER INCHARGE	Date Received	: 21-Jun-2022
Address	: Community Health Centre, T.V.Malai Road, Mannadipet, Puducherry - 605501.	Date Calibrated	: 21-Jun-2022
Instrument	: REFRIGERATOR	Due Date	: 20-Jun-2023
Manufacturer	: WHIRLPOOL	Ambient Temp.	: 26.3±3°C
Model No.	: GENIUS	Relative Humidity	: 61.4±15% R.H
Part No.	: ---	Serial No.	: ---
Location	: 37A/LAB PACS	Instrument ID.	: RF-08
		Range (Res.)	: ---
		Accuracy / MPE	: ---
		Test Results	: As Found

METHOD OF CALIBRATION


1. The instrument was calibrated as per method no. TEC at customer premises in accordance with ISO/IEC 17025:2017(E).

Master Instrument Used	Cert. No	Cal. Date	Validity	Traceability
1. Temperature Sensor (PT-100)	T021422	04-Feb-2022	03-Feb-2023	NLI
2. Multifunction Process Calibrator	E 1022/2022	21-May-2022	20-May-2023	NLI


RESULTS OF THERMAL CALIBRATION

INSTRUMENT	MASTER	TOLERANCE	STATUS	MEASUREMENT
SETTING(°C)	READING(°C)	(±)°C		UNCERTAINTY(±)°C
TOP	-14.9	---	---	0.59
MIDDLE	2.1	---	---	0.59
BOTTOM	2.3	---	---	0.59

REMARKS / DEVIATIONS IF ANY (SEE BACK OF THIS PAGE)


S. Gunasekaran
Calibration Officer




A. Murugappan / L. Muthuraman / S. Preetha
Authorised Signatory
NorthLab (I) Pvt Ltd

X:\Cert-2022\NABL-Cert\Temp-SiteA-TCOMHON HEALTH\2360622-T208\96E.XLS\96E1

End of Certificate



“களவினால் ஆகிய ஆக்கம் அளவிறந்து
ஆவது போலக் கெடும்” -குறள் 283

The property, which is acquired by fraud, will pass all bounds
and perish, even while it seems to increase. - Kural 283