



Scientific Temper



Honesty, Humility

**NORTHLAB (INDIA) PVT. LTD.**

Regd. Off. : Plot No.21, Phase II, Hosur Valley Site, Viswanathapuram, Hosur - 635 109, India  
 Ph.: +91 9176651123 E-mail: saleshosur@northlab.in, Website: www.northlab.in  
 Mechanical Lab : Plot No.171, 9th Cross Tirupathi Majestic, Viswanathapuram, Hosur - 635 109

CC-2561

**NORTHLAB****Calibration Certificate**

Cert. No.	: 1421022-V01	ULR NO	: CC256122100006001F
Date of Issue	: 14-Oct-2022	Page	: 1 of 1
Customer	: THE MEDICAL OFFICER INCHARGE	Date Received	: 13-Oct-2022
Address	: Community Health Centre, T.V.Malai Road, Mannadipet, Puducherry - 605501.	Date Calibrated	: 14-Oct-2022
		Due Date	: 13-Oct-2023
		Ambient Temp.	: 25±1°C
		Relative Humidity	: 55±5% R.H
Instrument	: <b>MICROPIPETTE</b>	Serial No.	: ---
Manufacturer	: RIGHT Q PLUS	Instrument ID.	: ---
Model No.	: ---	Range (Res.)	: 10-100 µl (1 µl)
Part No.	: ---	Accuracy / MPE	: ± 0.12-0.8 µl
Location	: ---	Test Results	: As Found

**METHOD OF CALIBRATION**

The instrument was calibrated as per method no. MVMP (Based on ISO : 8655-6:2002) at Northlab in accordance with ISO/IEC 17025:2017(E).

Master Instrument Used	Cert. No	Cal. Date	Validity	Traceability
1. Electronic Weighing Balance	NLHWB220G/22	23-May-2022	22-May-2023	NLI

**RESULTS OF MECHANICAL CALIBRATION**

INSTRUMENT READING (µl)	MEASURED VALUE (µl)	SYSTEMATIC ERROR (µl)	RANDOM ERROR (µl)	SYS. ERROR TOLERANCE (±) (µl)	STATUS	MEASUREMENT UNCERTAINTY ± (µl)
10	10.0701	0.0701	0.0000	0.12	PASS	1.56
50	50.4830	0.4830	0.0001	0.50	PASS	1.56
100	100.6892	0.6892	0.0001	0.80	PASS	1.56

**REMARKS / DEVIATIONS IF ANY (SEE BACK OF THIS PAGE)**

- The instrument is within the Maximum Permissible Error.

B. Muthukumar  
Calibration Officer



A. Murugappan / A. Balamurugan  
Authorised Signatory  
NorthLab (I) Pvt. Ltd

\*\*\* End of Certificate \*\*\*

“களவினால் ஆகிய ஆக்கம் அளவிறந்து  
உலகப் போலக் கெடும்” -குறள் 283



The property, which is acquired by fraud, will pass all bounds and perish, even while it seems to increase. - Kural 283