

VIEW LAB MONTHLY SUMMARY

Lab Name	TELANGANA DIAGNOSTIC HUB		Details About Robust Analysis							
Lab No	14984		Detail About Monthly Summary							
Month	March		Detail about SDI							
Year	2023									
Constituent Group	Chemistry I									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 21/03/2023							
			Date of Report Published : 04/04/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE	Hexokinase	Any Analyser (Automation / Semi Automation)	190	330.16	5.20	17.16	330 mg/dL	-0.01	2.49
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	700	58.51	9.15	5.35	33 mg/dL	-4.77	0.40
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	795	1.84	14.24	0.26	2.03 mg/dL	0.73	0.02
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	927	5.66	12.32	0.70	6.2 mg/dL	0.77	0.05
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	1082	5.49	10.00	0.55	5.3 g/dL	-0.35	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	751	3.31	7.29	0.24	3.1 g/dL	-0.87	0.02
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	909	9.37	7.83	0.73	8.5 mg/dL	-1.19	0.05
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	931	4.50	17.14	0.77	3.8 mg/dL	-0.91	0.05
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	1038	110.71	10.15	11.24	106 mg/dL	-0.42	0.70
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	1004	146.11	8.05	11.77	157 mg/dL	0.93	0.74
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	701	28.43	19.18	5.45	30 mg/dL	0.29	0.41
12	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1032	69.96	12.49	8.74	60 U/L	-1.14	0.54
13	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	949	197.38	18.40	36.32	193 U/L	-0.12	2.36
14	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	713	52.08	20.82	10.84	53 U/L	0.08	0.81
15	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	480	106.12	17.92	19.02	277 U/L	8.98	1.74
SDI Range				Interpretation						

Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out