



ARCHERCAL Private Limited

(Formerly known as Archerchem Calibrations Pvt Ltd)

Laboratory : Unit No. B/103, Tanvi's Tanishka Industrial Estate, Tanvi Complex,
Near HP Petrol Pump, Dahisar (East), Mumbai - 400 068. India. • Tel.: +91 22 20892984
E-mail : sales1@archercal.com / admin@archercal.com • Website : www.archercal.com

CIN : U74999MH2018PTC304510



CALIBRATION CERTIFICATE

7.8/R/M&V-01, Rev.00

Certificate No.: ACPL/MV/L/0602/02/23-24

ULR No.: CC230123000004375F

Date of Calibration:	06-04-2023	Issue Date:	07-04-2023
Next Recommended Calibration Due Date:	05-04-2024	Page:	01 of 02

CUSTOMER NAME & ADDRESS : THE LAB BEYOND EXCELLENCE
A2, 001/002, Prabhakar CHS LTD, Station Rd, Sector 4, Shanti Nagar,
Mira Road East, Mumbai, Maharashtra 401107

DATE OF RECEIPT : 05-04-2023

CALIBRATED AT : Lab

INSTRUMENT : Micropipette

ID. NO. : TLP/MP/02

CONDITION OF ITEM : Good

CALIBRATION PROCEDURE : WI/APL/CAL/MV/04
ISO 8655-6 (Latest Edition)

ENVIRONMENTAL CONDITION :

- Temperature : 23 ± 1 °C
- Air Pressure : 1010 ± 30 hPa
- Relative Humidity : 50 ± 10 % RH
- Water Temperature: 23.4 °C
- Z Correction Factor (µl/mg) 1.00360
- Y Correction Factor ((1/°C)x°C) 1.00162

TRACEABILITY : This Certificate is issued in the field of calibration and provides traceability of measurement results to International systems of units (SI)

CALIBRATIONS RESULTS : The results have been presented on pages(s) 2 of this certificate including uncertainty of measurements.

UNCERTAINTY OF MEASUREMENT : The uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 corresponds to confidence level of 95%.

CONFORMITY STATEMENT : On the Basis of Calibration results, it has been found that instruments submitted for calibration meets the requirements specified in standard 8655

REMARKS :

- The measured values mentioned are the average of 10 readings
- The Reported Volume at 27°C is by the formula,
 $V_{27^{\circ}\text{C}}(\mu\text{l}) = m(\text{mg}) \times Z(\mu\text{l}/\text{mg}) \times Y((1/^{\circ}\text{C}) \times ^{\circ}\text{C})$.
- The test liquid used for calibration is distilled water as per ISO3696 requirement.

Calibrated By:

Sapana Vaidya

Calibration Engineer



Approved By :

Kailas Chilap

Technical Director



ARCHERCAL Private Limited

(Formerly known as Archerchem Calibrations Pvt Ltd)

Laboratory : Unit No. B/103, Tanvi's Tanishka Industrial Estate, Tanvi Complex,
Near HP Petrol Pump, Dahisar (East), Mumbai - 400 068. India. • Tel.: +91 22 20892984
E-mail : sales1@archercal.com / admin@archercal.com • Website : www.archercal.com
CIN : U74999MH2018PTC304510



CALIBRATION CERTIFICATE

7.8/R/M&V-01 Rev.00

Certificate No.: ACPL/MVL/0602/02/23-24

ULR No.: CC230123000004375F

Date of Calibration:	06-04-2023	Issue Date:	07-04-2023
Next Recommended Calibration Due Date:	05-04-2024	Page:	02 of 02

Micropipette

Make :	VTTALPETTE	Range :	5 - 50 μ l
Model :	---	Least Count :	0.1 μ l
Sr.No. :	12104334	Accuracy :	As per ISO 8655-2
Tag/Id.No. :	TLP/MP/02	Location	----

Discipline : Mechanical Calibration

Product Group : Mass & Volume - Volume

Measurement Result :

Sr. No.	Nominal Volume in μ l	1	2	3	4	5	6	7	8	9	10	Volume at 27°C (Average) in μ l
		Mass of Water in mg										
1	5	5.091	5.129	5.121	5.089	5.083	5.071	5.083	5.041	5.069	5.072	5.111
2	25	25.103	25.181	25.171	25.143	25.141	25.122	25.143	25.137	25.183	25.163	25.280
3	50	50.091	50.211	50.141	50.131	50.123	50.149	50.143	50.149	50.141	50.137	50.404

Error (A) in μ l	Accuracy ($\pm A$) as per ISO 8655 in μ l	Precision (σ) in μ l	Precision ($\pm CV$) as per ISO 8655 in μ l	Expanded Uncertainty (\pm) in μ l	Status
0.111	0.50	0.026	0.20	0.015	Within Accuracy
0.280	0.50	0.026	0.20	0.015	Within Accuracy
0.404	0.50	0.030	0.20	0.015	Within Accuracy

REFERENCE STANDARDS USED

DESCRIPTION & MAKE	RANGE	ASSUMED DENSITY	Sr.No./ ID.NO.	TRACEABILITY	VALID UPTO
Digital Weighing Balance, RADWAG	Max Capacity : 0 - 5.1 g	--	513955/ ACPL/MUM/M&V/01	ACPL MUMBAI (ACPL/M*/S/0736/01/ 22-23)	02-06-2023

NOTE

- : The report refer only the particular item calibrated at site / laboratory.
- : The calibration result reported are valid at the time of and under the condition of measurement
- : Certificate should not be reproduced, except in full without the prior permission of Laboratory
- : Any correction in this certificate invalidates the certificate
- : The calibration of under test is meant for scientific and industrial purpose only.

Calibrated By: Sapana Valdya Calibration Engineer		Approved By : Kailas Chilap Technical Director
---	--	--

***** End of Certificate *****