



National AIDS Control Organisation
India's Voice against AIDS
Ministry of Health & Family Welfare, Government of India
www.naco.gov.in

PROFICIENCY TESTING SCHEME FOR HIV SEROLOGY

JIPMER, Puducherry
JAWAHARLAL INSTITUTE OF POSTGRADUATE MEDICAL EDUCATION AND RESEARCH (JIPMER),
DEPARTMENT OF MICROBIOLOGY, DHANVANTARI NAGAR, GORIMEDU, 605006

Report released: **02-02-2023 01:37:58 PM**

Proficiency Testing (PT) Report

Name of PT Panel Provider (NRL): **TNMGRU, Chennai, Tamil Nadu**
Name of Panel distributing Laboratory (SRL): **JIPMER, Puducherry**

Participant Lab: **ICTC-CHC-TNR**
Laboratory ID: **3400076**
PT round: **SECOND - 2022-23**

Panel issued: **2023/01/19**
Panel received: **2023-01-19**
Panel tested: **2023-01-19**

PARTICIPANT PERFORMANCE REPORT

Panel Member	Panel ID	Results of Participating Laboratory	Final Status	Laboratory Performance
Panel Member - 1	1	Positive	Concordant	Satisfactory
Panel Member - 2	2	Negative	Concordant	
Panel Member - 3	3	Negative	Concordant	
Panel Member - 4	4	Positive	Concordant	

"This report is confidential and the performance is shared with the participating laboratory, respective National Reference Laboratory, Apex laboratory and NACO only".
Kindly contact SRL In-charge for any queries.

[Signature]
MEDICAL OFFICER INCHARGE
ICTC, THIRUNALLAR
KARAIKAI

Report authorized by
Dr. Rakesh Singh- SRL In-Charge

-: End of report :-



JAWAHARLAL INSTITUTE OF POSTGRADUATE
MEDICAL EDUCATION AND RESEARCH
STATE REFERENCE LABORATORY FOR HIV
DEPARTMENT OF MICROBIOLOGY



External Quality Assessment Scheme for HIV Serology

JIP/SRL/HIV/23- 138

Date: 27.01.2023

EQAS REPORT (SRL to ICTCs/Blood banks)

Round: II

Name of Proficiency Testing Provider: SRL, JIPMER

Date Panel Distributed: 19.01.2023

Name of the participating laboratory: ICTC, CHC, Thirunallar

Date results received: 27.01.2023

Sample Code	Standard Result	Results obtained by participating laboratory	% Matching
R2/22-23/1	R	R	100 %
R2/22-23/2	NR	NR	100 %
R2/22-23/3	NR	NR	100 %
R2/22-23/4	R	R	100 %

Keyword: R: Reactive, NR: Non reactive

Remarks: Performance is Satisfactory

Name of Laboratory In-charge: Dr.Rakesh Singh

Signature of Laboratory In-charge

MEDICAL OFFICER INCHARGE
ICTC, THIRUNALLAR,
KARAIKAL