



Scientific Temper



Honesty, Humility



CC-2457

**NORTHLAB (INDIA) PVT. LTD.**Laboratory : Plot No. 1, S.V Nagar, Perumalpattu, Thiruvallur - 602 024.
Ph.: +91 89399 23940 / 89399 40050 E-mail: sales@northlab.in Website: www.northlab.in**Calibration Certificate**

Cert. No. : 2360622-T212	ULR NO : CC245722900002543F
Date of Issue : 23-Jun-2022	Page : 1 of 1
Customer : ICTC	Date Received : 22-Jun-2022
Address : Community Health Centre, Thenoor Main Road, Thenoor - 609 607, Thirunallar Post, Karaikkal.	Date Calibrated : 22-Jun-2022
	Due Date : 21-Jun-2023
	Ambient Temp. : 26.3±3°C
	Relative Humidity : 61.4±15% R.H
Instrument : REFRIGERATOR	Serial No. : HRD 262
Manufacturer : HAIER	Instrument ID. : ---
Model No. : HRD-2623/2021	Range (Res.) : ---
Part No. : ---	Accuracy / MPE : ---
Location : ICTC-LAB	Test Results : As Found

METHOD OF CALIBRATION


1. The instrument was calibrated as per method no. TEC at customer premises in accordance with ISO/IEC 17025:2017(E).

Master Instrument Used	Cert. No	Cal. Date	Validity	Traceability
1. Temperature Sensor (PT-100)	T021422	04-Feb-2022	03-Feb-2023	NLI
2. Multifunction Process Calibrator	E 1022/2022	21-May-2022	20-May-2023	NLI

RESULTS OF THERMAL CALIBRATION


INSTRUMENT	MASTER	TOLERANCE	STATUS	MEASUREMENT
SETTING(°C)	READING(°C)	(±)°C		UNCERTAINTY(±)°C
TOP	-13.6	---	---	0.59
MIDDLE	2.6	---	---	0.59
BOTTOM	2.8	---	---	0.59

REMARKS / DEVIATIONS IF ANY (SEE BACK OF THIS PAGE)


S. Gunasekaran
Calibration Officer

X:\Cert-2022\NABL-Cert\Temp-Site\A~1\COMMOM HEALTH\2360622-T212\96E.XLS\96E1




A. Murugappan / L. Muthuraman / S. Preetha
Authorised Signatory
NorthLab (I) Pvt. Ltd

End of Certificate



“களவினால் ஆகிய ஆக்கம் அளவிறந்து
ஆவது போலக் கெடும்” –குறள் 283

The property, which is acquired by fraud, will pass all bounds
and perish, even while it seems to increase. - Kural 283