



Scientific Temper



Honesty, Humility



CC-2457

**NORTHLAB (INDIA) PVT. LTD.**Laboratory : Plot No. 1, S.V Nagar, Perumalpattu, Thiruvallur - 602 024.
Ph.: +91 89399 23940 / 89399 40050 E-mail: sales@northlab.in Website: www.northlab.in**Calibration Certificate**

Cert. No.	: 2870622-T05	ULR NO	: CC245722400001303F
Date of Issue	: 24-Jun-2022	Page	: 1 of 1
Customer	: ICTC	Date Received	: 23-Jun-2022
Address	: Community Health Centre, Thenoor Main Road, Thenoor - 609 607, Thirunallar Post, Karaikkal.	Date Calibrated	: 24-Jun-2022
		Due Date	: 23-Jun-2023
		Ambient Temp.	: 23±3°C
		Relative Humidity	: 60±15% R.H
Instrument	: DIGITAL THERMOMETER WITH SENSOR	Serial No.	: 201506012119
Manufacturer	: III INSTRUMENTICS	Instrument ID.	: ---
Model No.	: ---	Range (Res.)	: -50~300°C(0.1°C)
Part No.	: ---	Accuracy / MPE	: ±1°C
Location	: ---	Test Results	: As Found

METHOD OF CALIBRATION

1. The instrument was calibrated as per method no. TTS at Northlab in accordance with ISO/IEC 17025:2017(E).

Master Instrument Used

1. Platinum Resistance Thermometer (PRT) With DMM

Cert. No

TSC/21-22/5622-2

Cal. Date

03-Aug-2021

Validity

02-Aug-2022

Traceability

TRANSCAL

RESULTS OF THERMAL CALIBRATION

MASTER READING (°C)	INSTRUMENT READING (°C)	CORRECTION (°C)	TOLERANCE (±) °C	STATUS	MEASUREMENT UNCERTAINTY (±) °C
-19.93	-19.7	-0.23	1.0	PASS	0.39
0.18	0.5	-0.32	1.0	PASS	0.39
50.23	50.0	0.23	1.0	PASS	0.39
150.04	149.7	0.34	1.0	PASS	0.39
249.96	249.4	0.56	1.0	PASS	0.39


REMARKS / DEVIATIONS IF ANY (SEE BACK OF THIS PAGE)

1. The instrument is within the Maximum Permissible Error.

D.Kamali
D.Kamali

Calibration Officer

Y:\Cert-2022\NABL-Cert\Temperature\~R\PONDICHERY AIDS\2870622-T05\96E1.XLS

A.Murugappan / L.Muthuraman / S.Pradha
Authorised Signatory
NorthLab (I) Pvt. Ltd

End of Certificate

“கள்ளினால் ஆகிய ஆக்கம் அளவிறந்து
ஆவது போலக் கெடும்” -குறள் 283The property, which is acquired by fraud, will pass all bounds
and perish, even while it seems to increase. - Kural 283