

Ortho Clinical Diagnostics

403, Leela Business Park, Andheri Kurla Road,
Andheri East, Mumbai - 400059
T : +91 22 6787 9300
F : +91 22 6787 9333

Calibration Certificate

The below mentioned instrument has been calibrated and tests performed to check the system performance.

Instrument : VITROS XT7600
Serial No : 7600 0983
Customer Name : USHAKAL ABHINAV SPECIALITY HOSPITAL, SANGLI
Calibration performed on : 28/01/2023

The system's calibration includes optics calibration and checking the reproducibility performance of the instrument as per the guidelines provided by the manufacturer.

Next Calibration will be performed in **JULY 2023**.

For **Ortho Clinical Diagnostics India Pvt Ltd.**



Onkar Gajanan Deshpande
Field Engineer - Ortho Care Service
Pune

Date: 28/01/2023