

Issued By: QA

Effective Date:
02.01.2018

Revision: A

FM-75-08-2
Page 1 of 1

Approved By: Dr. S. Jacob Inbaneson

Customer Name: Vk medical and scan center

Address: Annangay - Chennai - 600102

Contact Person: Mr. Kathiravan

Telephone No: 9840883350

Email ID:

INSTRUMENT DETAILS

Model Number: ADx - 200

Date of Installation: 15.02.2022

Serial Number:

Date of Visit: 17.02.2023

Warranty Information: Under Warranty

Out of Warranty

Under AMC

NATURE OF COMPLAINT

Pms, and optocoupler error

PROBLEM FACED DURING VISIT

Mixedly harm optocoupler & normal

ACTION TAKEN

Clear all optocoupler sensor,
and replace pms Done now Disturb
warranty bond Range (34-65) Comp

FEEDBACK / COMMENTS

good.

Engineer Signature: [Signature]

Customer Signature: [Signature]

Engineer Name: R. PRABU

Customer Name: V. Kathiravan

Date: 17.02.2023

Date: 17/2/2023

Customer Stamp:





FULLY-AUTO CHEMISTRY ANALYZER

ADX-AUTOCHEM-200