

- B. Waste pump
- C. ISE sensor system

5.Heater / Cooler

- A. Preheat temperature
- B. Reaction temperature
- C. Cooler temperature

6.cover / Door Latches

- A. Reagent cover
- B. Cuvette cover
- C. Drawer

7.System security test

Auto print

- A. Auto print all sample result
- B. Auto print all patient result
- C. Header report
- D. Patient Report test order

- A. Operator ID
- Position for stat
- Debug mode
- Auto run after inventory
- Calibration Result
- QC Result
- Worklist warning

CERTIFICATED BY



A.Ruban Harish

Service Engineer

NOTE:

- A. This certificate refers only to the particular item submitted for calibration.
- B. The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.


CALIBRATION CERTIFICATE

CUSTOMER NAME : Bio 5 ira Labs
ADDRESS : Mangalam Plaza, Door no: 226/5A, MGR Main Rd, Kandhanchavadi,
Perungudi, Chennai, Tamil Nadu 600096
CONTACT PERSON : Dr.Madhan
INSTRUMENT : FULLY AUTO BIOCHEMISTRY ANALYSER.
MODEL : EASY RA
S/NO : 20721031718
CALIBRATED DATE : 20/04/2023
CALIBRATION DUE DATE: 19/04/2024

THE FOLLOWING SUBSYSTEM HAS BEEN CALIBRATED IN THE INSTRUMENT.

- 1.REAGENT/SAMPLE AREA
- 2.REACTION AREA
- 3.TRANSFER ARM / PROBE
- 4.FLUID DRAWER
- 5.COOLER/HEATER
- 6.COVER / DOOR LATCH
- 7.PRECISION TEST

CERTIFICATED BY



A.Ruban Harish

Service Engineer