

**Date: 03/02/2023**

### **CALIBRATION CERTIFICATE**

This is to certify that the below-mentioned Instrument was calibrated on 03/02/2023 at **BIO 5 LAB, Mangalam Plaza, Door no: 226/5A, MGR Main Rd, Kandhanchavadi, Perungudi, Chennai, Tamil Nadu 600096.** And found working satisfactorily.


Instrument Name : Semi Auto Biochemistry Analyzer  
Model : YUMIZEN CA-60  
Serial No : 3130

Calibration Done on : 03/02/2023  
Next Calibration Due on : 02/02/2024

#### **CALIBRATION PARAMETERS:**

1. Input Power - Ok
  - i. Ph To N -229 Volts
  - ii. E To Ph -232 Volts
  - iii. N To E - 03 Volts
2. Display Test - Ok
3. Touch Screen Test - Ok
4. Pump Calibration Test -Ok
5. Optical Calibration Test - Ok
6. Printer Test - Ok

For Horiba India Pvt Ltd,



Team Lead Customer Support Engineer,  
P Tamilarasan