



VIEW LAB MONTHLY SUMMARY

Lab Name	CITYLAB	Details About Robust Analysis
Lab No	2156	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 16/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1561	308.70	9.98	30.81	299 mg/dL	-0.31	1.56
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1071	57.38	12.47	7.16	55 mg/dL	-0.33	0.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1156	1.86	11.88	0.22	1.62 mg/dL	-1.09	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1367	5.35	15.48	0.83	4.96 mg/dL	-0.47	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1230	5.53	10.94	0.60	4.72 g/dL	-1.34	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1198	3.31	9.84	0.33	3.5 g/dL	0.58	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1224	4.62	19.05	0.88	3.91 mg/dL	-0.81	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1526	112.67	12.38	13.95	95 mg/dL	-1.27	0.71
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1378	140.75	14.47	20.37	173 mg/dL	1.58	1.10

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links