

CALIBRATION CERTIFICATE

NAME OF THE CUSTOMER Sri Sai healthcare unit Sy no 83/2nd floor, Suchitra, Jeedimetla Hyderabad, Telangana, 500067.	MAKE Sysmex Corporation Sysmex India Pvt. Ltd. 1002, Damji Shamji Business Galleria, 10th Floor, LBS Marg Kanjurmarg (West), Mumbai 400 078, India Tel: +91 (22) 6112 6666 Fax: +91 (22) 2577 6790
PRODUCT NAME XN-350 HEAMATOLOGY ANALYZER	SERIAL NO 16447

This is to certify that the XN-350 Hematology Analyzer has been calibrated with XN CAL
 Calibrator (Lot No 23322101, Exp. Date 01/01/2023)

S.NO	TEST PARAMETERS	MEASURED/CALIBRATED	REMARKS
1.	Start up	WBC/RBC/HGB/PLT	PASS
2.	Reproducibility	WBC/RBC/HGB/MCV/HCT/PL T	PASS
3.	QC	WBC/RBC/HGB/HCT/PLT	PASS
4.	Calibration (Lot No. 23322101)	WBC/RBC/HGB/HCT/PLT	PASS

Date of Calibration: 17/12/2022	Due date for Calibration: 16/12/2023
---------------------------------	--------------------------------------

NAME OF THE ENGINEER

G S Manikantha

Faias
SIGNATURE

G S Manikantha



Sysmex India Pvt Ltd
 HO. 1002, Damji Shamji Business Galleria, 10th Floor, LBS Marg, Kanjurmarg (West), Mumbai 400078, India
 Tel. +91-22-6112-6666 Fax. +91-22-2577-6790
 Factory: Village Malpur, Nalagarh Road, Baddi 173205, H. P. Tel. +91-9218422282/9816672282

www.sysmex.co.in

CIN: U33120MH1998PT115943

