



**Abbott**

Alere Medical Pvt. Ltd.  
No. 404, BPTP Park Centra,  
NH - 8, Sector - 30  
Opp. 32nd Milestone, Gurugram  
Haryana 122001, India

T: +91 124 4569000  
F: +91 124 4569011

**Calibration Certificate**

**Instrument Name** : Urometer120  
**Serial Number** : 12076629  
  
**Customer Name** : Tesla Diagnostics Center  
**Address** : Suchitra, Hyderabad  
  
**Standard used** : Self Calibration Mode  
**Standard Traceability** : In House Certificate

**Calibration**

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of  $333 \pm 10$ .

**QC Performance** : Acceptable  
**Self Calibration** : Acceptable

**Performance Qualification : Passed**

**Calibration Date** : 08.10.2022  
**Calibration** : Passed  
**Calibration Valid up to** : 07.10.2023

**For Alere Medical Pvt Ltd.**

V. U. R. T.



Upendrakumar V  
Engineer Technical Services  
M: 9052723178

REGISTERED OFFICE: 221 & 222, DLF TOWERS, SHIVAJI MARG, NAJAFGARH ROAD, NEW DELHI - 110015 (INDIA)  
TEL: +91-11-45089400  
CIN NO: - U74899DL1995PTC069456

alere.com



Scanned with OKEN Scanner