



April, 28' 2023

***Calibration Certificate for Semi Automated Biochemistry Analyser,
Model: EC5+V2 (Sr. No.: N130145)***

This is to certify that the Semi Automated Biochemistry Analyser, Model: EC5+V2, Sl. No. N130145 installed at Regional Research Institute of Unani Medicine, Bhadrak, Matha Sahi, Near NH 5, Odisha-756100, has been calibrated and the calibration value and all the required filters values are well within limit.

Pump Calibration Value: 2299(Range 2000-2500)

Filter Gain(Range 120000-170000)

- 1. Filter 340: 147016*
- 2. Filter 405: 147333*
- 3. Filter 450: 150993*
- 4. Filter 505: 153798*
- 5. Filter 546: 155089*
- 6. Filter 578: 144131*
- 7. Filter 630: 160629*
- 8. Filter 670: 160791*

This calibration is valid till April, 27' 2024.

For, Transasia Bio-Medicals Ltd.



Shibbananda Biswas

Regional Service Manager-Odisha