

**CALIBRATION REPORT**

ADL/BO/MS/C/0310/22-23

Customer: Radvision Diagnostics & Labs, Bangalore

Serial No.: 212130157299

Model : Mispa Viva

DOI: 21/09/2021

Date of Calibration: 25/Jan/2023

INSTRUMENT CONTROL

Visual Control: PASSED

Zero Drift after 1 Hour of zero setting (WL=340nm) 0.001

PHOTOMETRIC LINEARITY

Solution	WL (nm)	100%	75%	50%	25%
Potassium dichromate	340nm	100.2	74.6	51.8	25.2
P-Nitrophenol	405nm	100.4	74.5	50.3	25.8
Cobaltous Ammonium Sulphate	505nm	100.8	76.9	50.1	24.7

CHEMISTRY CONTROL

	TEST	REAGENT	CONTROL SERUM	THEORETICAL VALUE	READ VALUE	RANGE
END POINT	Glu		Lot No:32090353 Expiry :31/08/24	80.0	81.6	74.52 - 85.5
KINETIC	GLUT		Lot No:32090353 Expiry :31/08/24	44.0	41.1	39 - 49
FIXED TIME	Urea		Lot No:32090353 Expiry :31/08/24	30.8	28.6	26.8 - 34.8

Validity: 6 months from the date of calibration.

For Agappe Diagnostics Ltd.,

Authorized Signatory

Authorized Signatory



Name: Femin Simon

Designation: Engineer Customer Service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattomattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7005 | Email: agappe@agappe.in

EQUIPMENT PLANT
K/508-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jansingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4390 8009 | Email: mumbaioffice@agappe.in

DELHI OFFICE
D5M 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in