

DETAILS OF CONTROL RUN

Instrument Model : Mispa Count X - 3 Part Hematology Analyzer
 Serial No : 2121100010194
 Date of Control Run : 2023 - 02 - 14
 Control Used : Hematology Control (Tri Level)
 Lot No : 32110386
 Expiry : 2023 - 03 - 05

The results of repeatability test are summarized in the following table. Refer the attached control sheets for details.

Sample No	WBC K/ μ L	RBC M/ μ L	HGB g/dL	HCT (%)	PLT K/ μ L
1	7.1	4.3	12.7	36.5	268
2	7.0	4.4	13.0	38.3	272
3	7.1	4.2	13.2	38.	265
Instrument Mean	7.0	4.3	12.9	37.6	268
Range	6.2 - 7.8	3.5 - 5.3	12.1 - 13.9	32.5 - 43.5	231 - 295
CV %	11.42	20.9	6.97	14.6	11.94

This is to certify that the above mentioned product has been calibrated for CBC parameters (WBC, RBC, HGB, HCT, PLT) according to the standard operating procedure given by Agappe.

The control run after calibration shows good repeatability and precision

Next Calibration due on 24/July/2023

For AGAPPE DIAGNOSTICS LTD.

Authorized Signatory



Name: Femin Simon

Designation: Engineer Customer Service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,
 Nettad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
 406, Merin Matrix, Plot No-10, Block-DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jaisingh Business Centre, 119,
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4200 0000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4000 0416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-6, 8th Floor, Red Cross Bhawan, No.26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2226 8288
 Email: bangaloreoffice@agappe.in