

VIEW LAB MONTHLY SUMMARY

Lab Name **CARE DIAGNOSTICS**

Lab No **14167**

Month **February**

Year **2023**

Constituent Group **Chemistry II**

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 15/02/2023

Date of Report Published : 08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	179	119.55	7.87	9.41	125 mg/dL	0.58	1.41
2	UREA II	Urease UV / GLDH II	Mindray	142	129.04	9.05	11.68	76 mg/dL	-4.54	1.96
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Mindray	110	4.74	13.83	0.66	5.98 mg/dL	1.89	0.13
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	163	1.47	17.09	0.25	1.47 mg/dL	0.00	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	162	4.70	9.59	0.45	5.21 g/dL	1.13	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	179	2.94	10.32	0.30	3.27 g/dL	1.09	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	162	7.07	8.12	0.57	4.89 mg/dL	-3.82	0.09
8	CHOLESTROL II	CHOD-PAP II	Mindray	178	98.73	9.06	8.94	91 mg/dL	-0.86	1.34
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	183	90.91	12.86	11.69	84 mg/dL	-0.59	1.73

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name **CARE DIAGNOSTICS**

Lab No **14167**

Month **March**

Year **2023**

Constituent Group **Chemistry II**

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 12/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	190	318.46	8.85	28.19	301 mg/dL	-0.62	4.09
2	UREA II	Urease UV / GLDH II	Mindray	153	58.52	7.41	4.34	48 mg/dL	-2.42	0.70
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Mindray	114	1.90	9.75	0.18	1.81 mg/dL	-0.49	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	171	5.47	10.93	0.60	4.5 mg/dL	-1.62	0.09
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	187	5.42	8.55	0.46	4.6 g/dL	-1.77	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	185	3.34	7.97	0.27	3 g/dL	-1.28	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	167	4.32	16.73	0.72	3 mg/dL	-1.83	0.11
8	CHOLESTROL II	CHOD-PAP II	Mindray	198	109.71	8.91	9.78	93 mg/dL	-1.71	1.39
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	184	144.87	8.97	13.00	147 mg/dL	0.16	1.92

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	CARE DIAGNOSTICS	Details About Robust Analysis
Lab No	14167	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 15/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	188	149.06	7.59	11.31	147 mg/dL	-0.18	1.65
2	UREA II	Urease UV / GLDH II	Mindray	158	37.86	9.49	3.59	36.5 mg/dL	-0.38	0.57
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Mindray	119	2.70	11.98	0.32	3.08 mg/dL	1.18	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	176	2.96	12.62	0.37	2.84 mg/dL	-0.32	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	187	5.52	8.24	0.46	5.23 g/dL	-0.64	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	190	3.40	8.05	0.27	3.32 g/dL	-0.29	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	176	5.49	11.31	0.62	5.3 mg/dL	-0.31	0.09
8	CHOLESTROL II	CHOD-PAP II	Mindray	187	118.34	7.13	8.44	120.2 mg/dL	0.22	1.23
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	178	111.30	9.75	10.86	132.6 mg/dL	1.96	1.63

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out