




Rapid Diagnostic Pvt. Ltd.



Calibrator Calibration

Instrument: 5 Part Haematology Analyzer		Model: BC -6200		Serial No.: TW-03000933		Date of Cal. 20.07.2022	
Mode: OV-WB		Calibrator Lot No. PLUS0722				Exp.Date:05.08.2022	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D	
TARGET	9.05	4.59	13.6	92.5	255	9.05	
1.	8.46	4.71	14.4	94.8	261	8.64	
2.	8.47	4.63	14.4	94.7	261	8.33	
3.	8.62	4.63	14.4	94.6	252	8.48	
4.	8.66	4.70	14.3	94.2	246	8.46	
5.	8.52	4.64	14.3	94.6	247	8.45	
6.	8.68	4.68	14.3	94.1	249	8.58	
Mean	8.87	4.66	14.4	94.5	253	8.49	
CV(%)	1.135	0.792	0.20	0.30	2.8	1.278	
New Cal. Factor (%)	103.12	124.48	96.51	101.54	101.98		
Old Cal. Factors (%)	103.98	91.38	94.63	101.27	93.10		
Old TC1 (%)						101.16	
New TC1 (%)						100.95	

For Rapid Diagnostic Pvt. Ltd.


G. Nagaraju
Service Engineer


N. Mahender Reddy
Zonal Service Manager




Calibration Certificate

Customer Name	Telangana Diagnostics VIKARABAD
Address	Area Hospital Asifabad , Asifabad-504293
Contact Person	Mr. ARJUN
Contact No.	9701785212
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000933
Instrument Status	Under Warranty
Calibration Done On	27.07.2022
Next Calibration Due On	26.07.2023

For Rapid Diagnostic Pvt. Ltd.



G. Nagaraju
Service Engineer



N. Mahender Reddy
Zonal Service Manager