

SIMCO CALIBRATION LABORATORY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27 Plot No.151, Street No.3, Teacher's Colony, East Marrepally, Secunderabad - 26, Telangana

Tel: 040 2773 2341/342, 2773 1510, Mob: 77299 91231/33, 98480 46524.

E-mail: simcohyd@gmail.com, Info@simco-india.com, www.simcocalibrationlaboratory.com



CC-2806

CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page 1 of 1

Certificate No. : SL2211TS0614-005	Issue Date : 06-12-2022
1. Customer Name & Address: M/S. Telangana Diagnostics, (TD Hub) emergency opposite Rims district hospital 504001 Adilabad	ULR - CC 2 8 0 6 2 2 0 0 0 0 7 4 7 5 F
	Reference Date : 29-11-2022
	Calibration Date : 02-12-2022 Calibration Due Date : 01-12-2023

2. Details of Unit Under Calibration:

Description : Cold Room	Location : Pathology
Make : Blue Star	Model : GSM-2020
Range : 2 to 8 °C	
Resolution : 0.1 °C	

3. Detail of Standard Instruments Used :

Instrument Used	SI / Id No	Valid up to	Certificate No.
Universal Data Logger With Sensors (8-Channel)	SL/SMT/DL/02	28-07-2023	SL2207TL0536-001

4. Environmental Conditions: Standard Temperature : (25±4)°C Relative Humidity : (50±20) % RH

5. Calibration Procedure: SOP-TL-01

6. Thermal Calibration: Temperature

7. Calibration Results:

S No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Expanded Uncertainty in (± °C)
1	4.5	5.0	0.5	0.9

8. Remarks:

- The instrument/equipment is in good condition and was calibrated at Site
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported approximately at 95.45% confidence level with coverage factor $k=2$.

M. C. Suresh
Calibrated by

N.V. Kameswara Rao
Technical Manager
Authorised Signatory